

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

Agency: FDLE

Date/Time: 02/18/2024 23:40:00

Data Download Date/Time: 02/22/2024 04:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000229

Agency: FDLE

Date/Time: 01/30/2024 19:37:43

Data Download Date/Time: 02/07/2024 11:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

Agency: LARGO PD

Date/Time: 02/05/2024 15:33:19

Data Download Date/Time: 02/28/2024 10:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: LARGO PD

Date/Time: 02/27/2024 16:59:44

Data Download Date/Time: 02/28/2024 10:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

Agency: SUMTER COUNTY SO

Date/Time: 02/28/2024 18:03:40

Data Download Date/Time: 02/28/2024 19:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

Agency: SUMTER CO SO

Date/Time: 02/28/2024 18:05:55

Data Download Date/Time: 02/28/2024 18:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 01/01/2024 02:55:43

Data Download Date/Time: 02/05/2024 12:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: TITUSVILLE PD

Date/Time: 01/01/2024 06:57:01

Data Download Date/Time: 02/05/2024 12:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 01/17/2024 10:42:51

Data Download Date/Time: 02/05/2024 12:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: TITUSVILLE PD

Date/Time: 02/05/2024 10:36:20

Data Download Date/Time: 02/05/2024 12:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 02/05/2024 11:29:30

Data Download Date/Time: 02/05/2024 12:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

Agency: CLAY COUNTY SO

Date/Time: 01/27/2024 10:34:12

Data Download Date/Time: 01/27/2024 12:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 02/25/2024 06:37:37

Data Download Date/Time: 02/26/2024 08:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

Agency: LARGO PD

Date/Time: 02/08/2024 08:48:07

Data Download Date/Time: 02/28/2024 09:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: LARGO PD

Date/Time: 02/19/2024 08:29:53

Data Download Date/Time: 02/28/2024 09:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

Agency: PUTNAM COUNTY SO

Date/Time: 01/29/2024 14:16:28

Data Download Date/Time: 01/30/2024 12:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001157

Agency: POLK CSO

Date/Time: 02/13/2024 13:45:00

Data Download Date/Time: 02/13/2024 14:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

Agency: DAYTONA BEACH SHORES

Date/Time: 01/10/2024 20:58:26

Data Download Date/Time: 01/10/2024 22:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: DAYTONA BEACH SHORES

Date/Time: 02/06/2024 03:31:56

Data Download Date/Time: 02/06/2024 04:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

Agency: DAYTONA BEACH SHORES

Date/Time: 02/11/2024 04:58:24

Data Download Date/Time: 03/13/2024 18:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

Agency: EDGEWATER PD

Date/Time: 02/05/2024 11:33:48

Data Download Date/Time: 03/25/2024 14:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

Agency: MONROE COUNTY S.O.

Date/Time: 01/25/2024 11:24:19

Data Download Date/Time: 01/25/2024 17:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

Agency: OKEECHOBEE PD

Date/Time: 01/30/2024 01:54:56

Data Download Date/Time: 01/30/2024 02:43

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

Agency: EUSTIS P.D.

Date/Time: 01/29/2024 04:47:32

Data Download Date/Time: 01/29/2024 05:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: EUSTIS P.D.

Date/Time: 02/28/2024 04:55:35

Data Download Date/Time: 02/28/2024 05:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency: FL HIGHWAY PATROL

Date/Time: 01/28/2024 17:32:30

Data Download Date/Time: 01/28/2024 20:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004771

Agency: GULF BREEZE PD

Date/Time: 02/17/2024 18:43:33

Data Download Date/Time: 02/17/2024 21:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

Agency: LAKELAND PD

Date/Time: 01/29/2024 14:13:55

Data Download Date/Time: 01/29/2024 17:24

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

Agency: CLAY COUNTY SO

Date/Time: 01/27/2024 10:17:52

Data Download Date/Time: 01/27/2024 11:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 02/25/2024 07:14:22

Data Download Date/Time: 02/26/2024 08:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

Agency: CLAY COUNTY SO

Date/Time: 02/25/2024 07:15:44

Data Download Date/Time: 02/26/2024 08:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

Agency: BOCA RATON PD

Date/Time: 01/20/2024 12:53:05

Data Download Date/Time: 01/20/2024 13:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: BOCA RATON PD

Date/Time: 02/16/2024 17:54:33

Data Download Date/Time: 02/16/2024 18:41

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006738

Agency: FHP TRAINING ACADEMY

Date/Time: 02/27/2024 15:10:33

Data Download Date/Time: 03/07/2024 10:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007171

Agency: MARION COUNTY SO

Date/Time: 01/10/2024 11:15:09

Data Download Date/Time: 01/10/2024 12:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007363

Agency: ST JOHNS SO

Date/Time: 01/05/2024 13:12:27

Data Download Date/Time: 02/01/2024 21:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail
