

776
②

RECEIVED

MAR 09 2017

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAFAYETTE CO SO
Time of Inspection: 12:35

Date of Inspection: 03/06/2017

Serial Number: 80-000776
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018 | 0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018 | 0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 17316080A2 Exp: 08/06/2108 |
|----------------------------|--|--|--|--|
| 0.000 | 0.049 | 0.079 | 0.199 | 0.081 |
| 0.000 | 0.049 | 0.081 | 0.202 | 0.081 |
| 0.000 | 0.051 | 0.081 | 0.203 | 0.081 |

Number of Simulators Used: 4

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MATTHEW MANDOLA

Signature and Printed Name

03/06/2017
Date