

RECEIVED

FEB 02 2017

# Florida Department of Law Enforcement Alcohol Testing Program

FDLE  
Alcohol Testing Progra

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SUWANNEE CO SO  
Time of Inspection: 08:56

Date of Inspection: 02/02/2017

Serial Number: 80-000780  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       |     | No |
| Alcohol Free Subject Test: 0.000            |     | No |
| Mouth Alcohol Test: Slope Not Met           |     | No |
| Interferent Detect Test: Interferent Detect |     | No |
| Diagnostic Check (Post-Inspection): OK      |     | No |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#: _____<br>Exp: _____ | 0.08g/210L Test (g/210L)<br>Lot#: _____<br>Exp: _____ | 0.20g/210L Test (g/210L)<br>Lot#: _____<br>Exp: _____ | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#: _____<br>Exp: _____ |
|----------------------------|---|---|---|--|
|                            |   |   |   |  |
|                            |   |   |   |  |
|                            |   |   |   |  |

Number of Simulators Used: \_\_\_\_\_

Remarks:

BYPASS AI TO OPERATE INSTRUMENT

*RUB BK*

*\* Compliance not determined \**

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAKE L SHANAHAN

Signature and Printed Name

02/02/2017  
Date