



RECEIVED

Florida Department of Law Enforcement AUG 25 2017  
Alcohol Testing Program

FDLE  
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HOLMES COUNTY SO  
Time of Inspection: 16:30

Date of Inspection: 07/20/2017

Serial Number: 80-000784  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |



| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201603D<br>Exp: 03/08/2018 | 0.08g/210L Test (g/210L)<br>Lot#:201611B<br>Exp: 11/15/2018 | 0.20g/210L Test (g/210L)<br>Lot#:201702B<br>Exp: 02/23/2019 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:32316080A2<br>Exp: 01/05/2019 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.047   | 0.078   | 0.200   | 0.079   |
| 0.000                      | 0.048   | 0.078   | 0.200   | 0.079   |
| 0.000                      | 0.048   | 0.078   | 0.201   | 0.079   |

Number of Simulators Used: 5

Remarks:  
NA

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 Signature and Printed Name JOANNA N BOWMAN / Susan Borge  
 Date 07/20/2017  


RECEIVED

AUG 25 2017

# Florida Department of Law Enforcement Alcohol Testing Program

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HOLMES COUNTY SO  
Time of Inspection: 14:46

Date of Inspection: 06/03/2017

Serial Number: 80-000784  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#: 201603D<br>Exp: 03/08/2018 | 0.08g/210L Test (g/210L)<br>Lot#: 201611B<br>Exp: 11/15/2018 | 0.20g/210L Test (g/210L)<br>Lot#: 201702B<br>Exp: 02/23/2019 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#: 32316080A2<br>Exp: 01/05/2019 |
|----------------------------|--|--|--|--|
| 0.000                      | 0.048  | 0.078  | 0.200  | 0.080  |
| 0.000                      | 0.047  | 0.079  | 0.201  | 0.080  |
| 0.000                      | 0.048  | 0.079  | 0.200  | 0.080  |

Number of Simulators Used: 5


Remarks:

A F / M A: TOO MUCH MA



The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



SUSAN M BARGE

Signature and Printed Name

06/03/2017  
Date

RECEIVED

JUL 26 2017

# Florida Department of Law Enforcement Alcohol Testing Program

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HOLMES COUNTY SO  
Time of Inspection: 16:30

Date of Inspection: 07/20/2017

Serial Number: 80-000784  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201603D<br>Exp: 03/08/2018 | 0.08g/210L Test (g/210L)<br>Lot#:201611B<br>Exp: 11/15/2018 | 0.20g/210L Test (g/210L)<br>Lot#:201702B<br>Exp: 02/23/2019 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:32316080A2<br>Exp: 01/05/2019 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.047   | 0.078   | 0.200   | 0.079   |
| 0.000                      | 0.048   | 0.078   | 0.200   | 0.079   |
| 0.000                      | 0.048   | 0.078   | 0.201   | 0.079   |

Number of Simulators Used: 5

Remarks:  
NA

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name JOANNA N BOWMAN

07/20/2017  
Date

*Susan Borge*  
*Tpr. Susan M Borge*

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HOLMES COUNTY SO  
Time of Inspection: 14:46

Date of Inspection: 06/03/2017

Serial Number: 80-000784  
Software: 8100.27

| Check or Test                               | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:201603D<br>Exp: 03/08/2018 | 0.08g/210L Test (g/210L)<br>Lot#:201611B<br>Exp: 11/15/2018 | 0.20g/210L Test (g/210L)<br>Lot#:201702B<br>Exp: 02/23/2019 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:32316080A2<br>Exp: 01/05/2019 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.048   | 0.078   | 0.200   | 0.080   |
| 0.000                      | 0.047   | 0.079   | 0.201   | 0.080   |
| 0.000                      | 0.048   | 0.079   | 0.200   | 0.080   |

Number of Simulators Used: 5


**Remarks:**

A F / M A: TOO MUCH MA



The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



SUSAN M BARGE

Signature and Printed Name

06/03/2017  
Date