

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HOLMES COUNTY SO  
Time of Inspection: 07:25

Date of Inspection: 10/03/2017

Serial Number: 80-000784  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: \_\_\_\_\_

**Remarks:**

SKIPPED AI TO OPERATE INSTRUMENT

*SP*

*N/A Not A compliance check*

The above instrument complies ( ☒ ) does not comply (     ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Patrick J Murphy*

PATRICK J MURPHY

Signature and Printed Name

10/03/2017  
Date

*10/5/17  
JL*

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**

AUG 29 2017

**FDLE**

## AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Pro

Agency: HOLMES COUNTY SO  
Time of Inspection: 08:11

Date of Inspection: 08/22/2017

Serial Number: 80-000784  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32316080A2 Exp: 01/05/2019
0.000	0.049	0.079	0.203	0.080
0.000	0.050	0.079	0.203	0.080
0.000	0.050	0.079	0.203	0.080

Number of Simulators Used: 5

Remarks:  
NA

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



SUSAN M BARGE  
Signature and Printed Name

08/22/2017  
Date

RECEIVED

# Florida Department of Law Enforcement

## Alcohol Testing Program

AUG 25 2017

FDLE

Alcohol Testing Program

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HOLMES COUNTY SO  
Time of Inspection: 16:30

Date of Inspection: 07/20/2017

Serial Number: 80-000784  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32316080A2 Exp: 01/05/2019
0.000	0.047	0.078	0.200	0.079
0.000	0.048	0.078	0.200	0.079
0.000	0.048	0.078	0.201	0.079

Number of Simulators Used: 5

Remarks:  
NA

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*[Signature]*

Signature and Printed Name

JOANNA N BOWMAN

*[Signature]*

07/20/2017  
Date

*[Signature]*

RECEIVED

AUG 25 2017

# Florida Department of Law Enforcement Alcohol Testing Program

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HOLMES COUNTY SO  
Time of Inspection: 14:46

Date of Inspection: 06/03/2017

Serial Number: 80-000784  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 32316080A2 Exp: 01/05/2019
0.000	0.048	0.078	0.200	0.080
0.000	0.047	0.079	0.201	0.080
0.000	0.048	0.079	0.200	0.080

Number of Simulators Used: 5


**Remarks:**

A F / M A: TOO MUCH MA



The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



SUSAN M BARGE

Signature and Printed Name

06/03/2017  
Date

RECEIVED

JUL 26 2017

# Florida Department of Law Enforcement

## Alcohol Testing Program

FDLE  
Alcohol Testing Program

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HOLMES COUNTY SO

Time of Inspection: 16:30

Date of Inspection: 07/20/2017

Serial Number: 80-000784

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32316080A2 Exp: 01/05/2019
0.000	0.047	0.078	0.200	0.079
0.000	0.048	0.078	0.200	0.079
0.000	0.048	0.078	0.201	0.079

Number of Simulators Used: 5

Remarks:

NA

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

JOANNA N BOWMAN

07/20/2017  
Date

/ Susan Borge  
Tr. Susan M Borge

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HOLMES COUNTY SO  
Time of Inspection: 14:46

Date of Inspection: 06/03/2017

Serial Number: 80-000784  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32316080A2 Exp: 01/05/2019
0.000	0.048	0.078	0.200	0.080
0.000	0.047	0.079	0.201	0.080
0.000	0.048	0.079	0.200	0.080

Number of Simulators Used: 5

Remarks:

A F / M A:TOO MUCH MA

*SM*

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Susan M Barge*

SUSAN M BARGE

Signature and Printed Name

06/03/2017  
Date