RECEIVED
JAN 11 2018
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO
Time of Inspection: 11:41 Date of Inspection: 12/19/2017
Serial Number: 80-000795
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17817080A2 Exp: 08/05/2019
0.000	0.050	0.080	0.199	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.051	0.081	0.201	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAMES P BAXLEY
Signature and Printed Name

12/19/2017
Date

FDLE/ATP Form 40 – March 2004

RECEIVED
NOV 28 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO
Time of Inspection: 11:32 Date of Inspection: 11/21/2017
Serial Number: 80-000795
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1781708A2 Exp: 08/05/2019
0.000	0.049	0.079	0.197	0.078
0.000	0.050	0.080	0.199	0.078
0.000	0.050	0.080	0.199	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAMES P BAXLEY
Signature and Printed Name

11/21/2017
Date

FDLE/ATP Form 40 – March 2004

RECEIVED
SEP 13 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO
Time of Inspection: 11:54
Date of Inspection: 09/06/2017
Serial Number: 80-000795
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201403F Exp: 03/25/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17817080A2 Exp: 08/05/2019
0.000	0.049	0.080	0.199	0.079
0.000	0.050	0.081	0.200	0.079
0.000	0.050	0.082	0.200	0.079

Number of Simulators Used: 5

Remarks:
A F / M A:NO MOUTH ALCOHOL.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAMES P BAXLEY
Signature and Printed Name

09/06/2017
Date

Florida Department of Law Enforcement
Alcohol Testing Program

RECEIVED
SEP 13 2017
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO
Time of Inspection: 22:17 Date of Inspection: 08/25/2017
Serial Number: 80-000795
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:20160305D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17817080A2 Exp: 08/05/2019
0.000	0.050	0.081	0.200	0.079
0.000	0.050	0.081	0.201	0.079
0.000	0.051	0.081	0.201	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAMES P BAXLEY
Signature and Printed Name

08/25/2017
Date

FDLE/ATP Form 40 – March 2004

RECEIVED
AUG 04 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO
Time of Inspection: 15:05 Date of Inspection: 07/26/2017 Serial Number: 80-000795
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:19615080A2 Exp: 09/05/2017
0.000	0.050	0.080	0.198	0.080
0.000	0.051	0.081	0.199	0.080
0.000	0.051	0.081	0.200	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

James P. Baxley
Signature and Printed Name JAMES P BAXLEY

07/26/2017
Date

Florida Department of Law Enforcement
Alcohol Testing Program

RECEIVED
JUN 02 2017
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO
Time of Inspection: 15:20 Date of Inspection: 05/31/2017 Serial Number: 80-000795
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		NO
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:19615080A2 Exp: 09/05/2017
0.000	0.051	0.083	0.203	0.078
0.000	0.051	0.084	0.202	0.078
0.000	0.051	0.083	0.202	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

James A. Newsome

JAMES A NEWSOME
Signature and Printed Name

05/31/2017
Date

FDLE/ATP Form 40 -- March 2004

RECEIVED
MAY 23 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO
Time of Inspection: 15:15 Date of Inspection: 04/27/2017
Serial Number: 80-000795
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: ~~03/08/2013~~ 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:19615080A2 Exp: 09/05/2017
0.000	0.051	0.084	0.204	0.078
0.000	0.051	0.084	0.203	0.078
0.000	0.051	0.083	0.202	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

James A. Jewson JAMES A. NEWSOME
Signature and Printed Name

04/27/2017
Date

Florida Department of Law Enforcement
Alcohol Testing Program

RECEIVED
MAY 08 2017
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO
Time of Inspection: 15:15
Date of Inspection: 04/27/2017
Serial Number: 80-000795
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: ~~03/08/2018~~ 03/08/2018 AN.5-2-17	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:19615080A2 Exp: 09/05/2017
0.000	0.051	0.084	0.204	0.078
0.000	0.051	0.084	0.203	0.078
0.000	0.051	0.083	0.202	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

James A. Newsome
Signature and Printed Name JAMES A NEWSOME

04/27/2017
Date

FDLE/ATP Form 40 – March 2004

RECEIVED
APR 03 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO
Time of Inspection: 14:44
Date of Inspection: 03/30/2017
Serial Number: 80-000795
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:19615080A2 Exp: 09/05/2017
0.000	0.050	0.081	0.203	0.078
0.000	0.051	0.081	0.204	0.077
0.000	0.051	0.082	0.205	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

James A. Newsome

JAMES A NEWSOME
Signature and Printed Name

03/30/2017
Date

FDLE/ATP Form 40 – March 2004