

Florida Department of Law Enforcement Alcohol Testing Program

810
B

RECEIVED

MAR 09 2017

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Testing Program

Agency: DIXIE COUNTY SO
Time of Inspection: 11:19

Date of Inspection: 03/06/2017

Serial Number: 80-000810
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:04216080A3 Exp: 03/05/2018
0.000	0.050	0.079	0.201	0.000 / 0.079
0.000	0.049	0.079	0.200	0.078 / 0.079
0.000	0.050	0.079	0.200	0.078 / 0.079

Number of Simulators Used: 4

Remarks:

08: Control Outside Tolerance AIR LEAK.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MATTHEW MANDOLA

Signature and Printed Name

03/06/2017
Date