

80-000812

RECEIVED

DEC 23 2017

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Time of Inspection: 09:16

Date of Inspection: 12/23/2017

Serial Number: 80-000812
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17316080A2 Exp: 08/05/2018
/ 0.000	0.048	0.078	0.194	0.081
/ 0.000	0.048	0.077	0.194	0.081
/ 0.000	0.047	0.078	0.194	0.081

Number of Simulators Used: 5

Remarks:

00: Ambient Fail. - HAND SANITIZER ODOR PRESENT DURING ALCOHOL FREE TEST. ODOR REMOVED AND ALCOHOL FREE TEST SUCCESSFUL.

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt B. A. 1193

BRAD R COLLITO

Signature and Printed Name

12/23/2017
Date

80-000812 DB

RECEIVED

NOV 24 2017

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Time of Inspection: 07:48

Date of Inspection: 11/24/2017

Serial Number: 80-000812
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17316080A2 Exp: 08/05/2018
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.080	0.198	0.080
0.000	0.050	0.080	0.198	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I, certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature of Brad R. Collito

BRAD R COLLITO

Signature and Printed Name

11/24/2017
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Time of Inspection: 19:47

Date of Inspection: 10/08/2017

Serial Number: 80-000812
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28515080A2 Exp: 11/05/2017
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.198	0.077
0.000	0.049	0.078	0.197	0.077

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt B C 1193

BRAD R COLLITO

Signature and Printed Name

10/08/2017
Date

80-000812 DB

RECEIVED

SEP 18 2017

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Time of Inspection: 19:32

Date of Inspection: 09/18/2017

Serial Number: 80-000812
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28515080A2 Exp: 11/05/2017
0.000	0.048	0.078	0.196	0.000 / 0.078
0.000	0.048	0.079	0.197	0.000 / 0.078
0.000	0.048	0.079	0.197	0.000 / 0.078

Number of Simulators Used: 5

Remarks:
DRY GAS HOSE DISCONNECTED FROM PREVIOUS TEST.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt B. Collito 1193

BRAD R COLLITO

Signature and Printed Name

09/18/2017
Date

80-000812 2/3

RECEIVED

AUG 22 2017

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Time of Inspection: 20:19

Date of Inspection: 08/22/2017

Serial Number: 80-000812
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17316080A2 Exp: 08/05/2018
0.000	0.049	0.080	0.197	0.081
0.000	0.049	0.080	0.198	0.082
0.000	0.049	0.080	0.198	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



BRAD R COLLITO

Signature and Printed Name

08/22/2017
Date

80-00812 *ADBS*

RECEIVED
JUL 19 2017

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Time of Inspection: 19:33

Date of Inspection: 07/19/2017

Serial Number: 80-000812
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17316080A2 Exp: 08/05/2018
0.000	0.049	0.079	0.198	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.080	0.199	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt B. C. 1193

BRAD R COLLITO

Signature and Printed Name

07/19/2017
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
JUN 13 2017
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Time of Inspection: 21:24

Date of Inspection: 06/13/2017

Serial Number: 80-000812
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17316080A2 Exp: 08/05/2018
0.000	0.049	0.079	0.197	0.080
0.000	0.049	0.080	0.197	0.081
0.000	0.050	0.080	0.197	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

BRAD R COLLITO

BRAD R COLLITO

Signature and Printed Name

06/13/2017
Date

80-000812 DB

RECEIVED

MAY 11 2017

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Time of Inspection: 17:22

Date of Inspection: 05/11/2017

Serial Number: 80-000812
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:12716080A7 Exp: 06/05/2018
0.000	0.048	0.079	0.198	0.082
0.000	0.048	0.080	0.199	0.082
0.000	0.049	0.080	0.200	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature
BR 1193

BRAD R COLLITO

Signature and Printed Name

05/11/2017
Date

80-000812

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
APR 13 2017
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Time of Inspection: 01:57

Date of Inspection: 04/13/2017

Serial Number: 80-000812
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:12716080A7 Exp: 06/05/2018
0.000	0.047	0.077	0.195	0.081
0.000	0.047	0.078	0.197	0.081
0.000	0.048	0.079	0.198	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

BRAD R. COLLITO

Signature and Printed Name

04/13/2017
Date

80-000812 RMB

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

MAR 26 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Time of Inspection: 18:34

Date of Inspection: 03/26/2017

Serial Number: 80-000812
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:12716080A7 Exp: 06/05/2018
0.000	0.047	0.080	0.194	0.081
0.000	0.049	0.081	0.198	0.081
0.000	0.048	0.080	0.199	0.081


Number of Simulators Used: 5

Remarks:

A F / M A: Range Exceeded.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



BRAD R COLLITO

Signature and Printed Name


03/26/2017
Date

80-000812 RAS

HERNANDO COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000812
03/25/2017
Software: 8100.27

RECEIVED
MAR 26 2017
FDLE
Alcohol Testing Program

Test	g/210L	Time
Air Blank	0.000	00:57
Control Test	0.194	00:57
Air Blank	0.000	00:58
Control Test	0.198	00:58
Air Blank	0.000	00:59
Control Test	0.199	01:00
Air Blank	0.000	01:00
Control Test	0.199	01:01
Air Blank	0.000	01:01
Control Test	0.199	01:02
Air Blank	0.000	01:03
Control Test	0.201	01:03
Air Blank	0.000	01:04
Control Test	0.201	01:05
Air Blank	0.000	01:05
Control Test	0.200	01:06
Air Blank	0.000	01:06
Control Test	0.200	01:07
Air Blank	0.000	01:08
Control Test	0.200	01:08
Air Blank	0.000	01:09
Control Test	0.200	01:10
Air Blank	0.000	01:10
Control Test	0.201	01:11
Air Blank	0.000	01:11
Control Test	0.202	01:12
Air Blank	0.000	01:13
Control Test	0.200	01:13
Air Blank	0.000	01:14
Control Test	0.202	01:14
Air Blank	0.000	01:15
Control Test	0.200	01:16
Air Blank	0.000	01:16
Control Test	0.200	01:17
Air Blank	0.000	01:18
Control Test	0.199	01:18
Air Blank	0.000	01:19
Control Test	0.199	01:19
Air Blank	0.000	01:20
Control Test	0.199	01:21
Air Blank	0.000	01:21
Control Test Stats		
Average	0.1996	
Std Dev	0.0017	
Rel Std Dev(%)	0.8487	



Operator's Signature

80-000812

RECEIVED

MAR 20 2017

FDLE Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Time of Inspection: 22:18

Date of Inspection: 03/15/2017

Serial Number: 80-000812
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:12716080A7 Exp: 06/05/2018
0.000	0.050	/ 0.079	/ 0.198	0.082
0.000	0.050	/ 0.081	/ 0.204	0.082
0.000	0.050	/ 0.081	/ 0.204	0.081

Number of Simulators Used: 5

Remarks:

A F / M A: RESIDUAL FROM PREVIOUS TEST. 08: RFI Detect. 20: RFI Detect.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt B. Collito

BRAD R COLLITO

Signature and Printed Name

03/15/2017
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
MAR 26 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Time of Inspection: 20:13

Date of Inspection: 03/15/2017

Serial Number: 80-000812
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:12716080A7 Exp: 06/05/2018
0.000	0.079 / 0.049	0.079 / 0.080	0.198	0.082 / 0.082
0.000	0.080 / 0.050	0.080 / 0.081	0.200	/ 0.081
0.000	0.080 / 0.050	RFI / 0.081	0.200	/ 0.081

Number of Simulators Used: 5

Remarks:

05: Control Outside Tolerance SIMULATORS SWITCHED. 08: RFI Detect. 08: RFI Detect.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature
Brad R Collito

BRAD R COLLITO

Signature and Printed Name

03/15/2017
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
FEB 15 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Time of Inspection: 11:40

Date of Inspection: 02/15/2017

Serial Number: 80-000812
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:12716080A7 Exp: 06/05/2018
0.000	0.048	0.077	0.198	0.081
0.000	0.049	0.078	0.200	0.081
0.000	0.049	0.078	0.200	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Brad R Collito 1193

BRAD R COLLITO

Signature and Printed Name

02/15/2017
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JAN 19 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.
Time of Inspection: 09:05

Date of Inspection: 01/19/2017

Serial Number: 80000812
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28515080A2 Exp: 11/05/2017
0.000	RFI / 0.049	0.078	0.197	0.080
0.000	/ 0.049	0.080	0.198	0.079
0.000	/ 0.049	0.080	0.199	0.079

Number of Simulators Used: 5

Remarks:

05: RFI Detect CELL PHONE IN BOOKING AREA.

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sgt B. C. 1193

BRAD R. COLLITO
Signature and Printed Name

01/19/2017
Date