

80-000823 ~~8003~~

RECEIVED

DEC 12 2017

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 02:59

Date of Inspection: 12/09/2017

Serial Number: 80-000823
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:12716080A7 Exp: 06/05/2018
0.000	0.047	0.075 / 0.078	0.196	0.080
0.000	0.048	0.074 / 0.078	0.198	0.080
0.000	0.048	0.074 / 0.079	0.198	0.079

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

C. J. ... 456

Signature and Printed Name

CHESSHER / PITCHER M. J. ... 1437

12/09/2017
Date

RECEIVED

NOV 14 2017

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 21:25

Date of Inspection: 11/03/2017

Serial Number: 80-000823
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201703A Exp: 03/07/2019	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1766638 Exp: 11/05/2017
0.000	0.046	0.079	0.193	0.079
0.000	0.048	0.079	0.194	0.080
0.000	0.048	0.079	0.195	0.080

Number of Simulators Used: 5

Remarks:
A F / M A: Range Exceeded TOO MUCH MOUTH ALCOHOL.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Jose Chaves 950 _____ CHESSHER / PITCHER *M/D/A 1937*
Signature and Printed Name

11/03/2017
Date

80-000823 RMB

RECEIVED

NOV 02 2017

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 20:49

Date of Inspection: 10/25/2017

Serial Number: 80-000823
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1766638 Exp: 11/05/2017
0.000	0.049	0.080	0.196	0.079
0.000	0.049	0.081	0.197	0.080
0.000	0.050	0.081	0.197	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

G. P. Chesser 950

Signature and Printed Name

CHESSHER / PITCHER

M. A. Pitcher

10/25/2017
Date

80-000823 *ms*

RECEIVED
SEP 27 2017

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 10:59

Date of Inspection: 09/27/2017

Serial Number: 80-000823
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1766638 Exp: 11/05/2017
0.000	0.048	0.079	0.195	0.078
0.000	0.049	0.080	0.196	0.078
0.000	0.049	0.080	0.196	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

M. Bowden
M. Bowden

BOWDEN / CHESSHER *[Signature]*
Signature and Printed Name

09/27/2017
Date

80-000823

RECEIVED

JUL 24 2017

FDLE Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 17:50

Date of Inspection: 07/21/2017

Serial Number: 80-000823
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1766638 Exp: 11/05/2017
0.000	0.048	0.080	0.196	0.079
0.000	0.048	0.080	0.197	0.079
0.000	0.049	0.080	0.197	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PITCHER / LUPPINO

 Signature and Printed Name

07/21/2017
Date

RECEIVED

MAY 23 2017

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO

Time of Inspection: 11:58

Date of Inspection: 05/22/2017

Serial Number: 80-000823

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1766638 Exp: 11/05/2017
0.000	0.048	0.080	0.196	0.079
0.000	0.049	0.080	0.197	0.079
0.000	0.049	0.080	0.197	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.




MATTHEW BOWDEN / SANDRA CHESSHER

Signature and Printed Name

05/22/2017
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

MAY 23 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 15:59

Date of Inspection: 04/27/2017

Serial Number: 80-000823
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1766638 Exp: 11/05/2017
0.000	0.046	0.078	0.191	0.079
0.000	0.048	0.079	0.195	0.079
0.000	0.049	0.080	0.196	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

M. BOWDEN / 1106
[Signature]

Signature and Printed Name

SANDRA W CHESSHER

[Signature]

04/27/2017
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
JUL 07 2017
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 10:21

Date of Inspection: 06/29/2017

Serial Number: 80-000823
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1766638 Exp: 11/05/2017
0.000	0.047	0.078	0.195	0.079
0.000	0.047	0.079	0.195	0.079
0.000	0.048	0.080	0.196	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

M. Bowden 1100

BOWDEN / CHESSHER

Signature and Printed Name

06/29/2017
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

APR 01 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 22:34

Date of Inspection: 03/31/2017

Serial Number: 80-000823
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1766638 Exp: 11/05/2017
0.000	0.049	0.079	0.196	0.080
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.198	0.080

Number of Simulators Used: 3

Remarks:

LCSO INTOX LOANED TO LPD

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

John G Sommersdorf Signature and Printed Name JOHN G SOMMERSDORF

03/31/2017
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

FEB 05 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO

Serial Number: 80-000823

Time of Inspection: 00:39

Date of Inspection: 02/05/2017

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1766638 Exp: 11/05/2017
0.000	0.048	0.077	0.193	0.080
0.000	0.049	0.079	0.195	0.080
0.000	0.050	0.079	0.196	0.081

Number of Simulators Used: 3

Remarks:

LAKE COUNTY INTOX ON LOAN TO LEEBURG PD

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

John G. Sommersdorf

JOHN G SOMMERSDORF

Signature and Printed Name

02/05/2017

Date

80-000823

RECEIVED

JAN 30 2017

FDLE Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 09:45

Date of Inspection: 01/30/2017

Serial Number: 80-000823
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1766638 Exp: 11/05/2017
0.000	0.048	0.078	0.194	0.080
0.000	0.048	0.078	0.194	0.080
0.000	0.049	0.079	0.194	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MATTHEW BOWDIN
M/B

SIGNATURE AND PRINTED NAME SANDRA W CHESSHER *Sandra Chessher* 950

01/30/2017
Date