

907
SR

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO
Time of Inspection: 11:45

Date of Inspection: 11/27/2017

Serial Number: 80-000907
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:256835 Exp: 01/11/2019
0.000	0.050	0.080	0.199	0.079
0.000	0.050	0.080	0.199	0.079
0.000	0.050	0.080	0.199	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



GLENN CLINE

Signature and Printed Name

11/27/2017
Date

907
SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO
Time of Inspection: 10:50

Date of Inspection: 05/22/2017

Serial Number: 80-000907
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201603D 201601F Exp: 03/08/2018 1/22/18	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 256835 Exp: 01/11/2019
0.000	0.049	0.077	0.196	0.080
0.000	0.049	0.078	0.196	0.080
0.000	0.049	0.078	0.197	0.080

Number of Simulators Used: 5

Remarks:

SGT SPURLOCK PRESENT

Accidentally entered lot # for 0.05 solution.
Actual used was correct lot # 201601F
All

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GLENN CLINE

Signature and Printed Name

05/22/2017

Date



Florida Department of Law Enforcement

AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Manatee County SO

Instrument Serial Number: 80-000907

AGENCY INSPECTION DATA REVIEW
Agency Inspector: Glenn Cline
Date of Inspection: 5/22/17
Time of Inspection: 10:50
Agency Inspection Discrepancy: [] Incomplete [] Untimely/Not Received [x] Erroneous Information
[] Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.
[] Lot Number [x] Expiration Date for 0.08g/210L [x] Alcohol Reference Solution [] Dry Gas Standard is [x] Incorrect [] Expired.
[] FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report - Intoxilyzer 8000. The [] REASON for repeating the following test(s); OR the [] Possible Cause and Corrective Action Taken on the following test(s) was not recorded:
[] Alcohol Free Subject Test [] Mouth Alcohol Test [] Alcohol Free Test [] Interferent Detect Test
[] 0.05 g/210L Test [] 0.08 g/210L Test [] 0.20 g/210L Test [] 0.08 g/210L Dry Gas Standard Test
[] FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector."
[] The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC.
[] The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use.
[] The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.
[] The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.
[] Other: _____

OTHER ELECTRONIC DATA REVIEW
[] Login Records Date: _____
[] Cylinder Change Records Date: _____
[] Control Test Records Date: _____
[] Diagnostic Check Records Date: _____
Comments: _____

CORRECTIVE ACTION
[x] Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by 6/22/17 (Date).
[] Provide a written explanation regarding the referenced item(s) to the Department Inspector by _____ (Date).
[] Upload the Agency Inspection(s).
[] Remove the instrument from evidentiary use until otherwise directed by the Department.
[] No action required
[] Other: _____

Signature of Shayla Platt
Signature of Alcohol Testing Program Staff Member

6/8/2017
Date

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

Agency MANATEE COUNTY SO	Date 05/22/2017	Type Agency Inspection	
Inspector CLINE, GLENN,	Time 10:50:48	Software 8100.27	
0.05 Lot # 201603D	0.08 Lot # 201603D	0.20 Lot # 201604C	0.08 Gas Lot # 256835
Expiration 03/08/2018	Expiration 03/08/2018	Expiration 04/05/2018	Expiration 01/11/2019

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

SGT SPURLOCK PRESENT

Data Download Date/Time 05/22/2017 10:59