80-000968 DOS

RECEIVED
DEC 02 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS
Time of Inspection: 20:46 Date of Inspection: 12/01/2017
Serial Number: 80-000968
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:163437 Exp: 10/30/2018
0.000	0.049	0.079	0.200	0.080
0.000	0.050	0.080	0.203	0.081
0.000	0.050	0.080	0.203	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D. Lyons

DANIEL E LYONS
Signature and Printed Name

12/01/2017
Date

FDLE/ATP Form 40 – March 2004

80-000968

RECEIVED
NOV 07 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS
Time of Inspection: 18:27 Date of Inspection: 11/07/2017
Serial Number: 80-000968
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:163437 Exp: 10/30/2018
0.000	0.049	0.080	0.198	0.079
0.000	0.049	0.080	0.199	0.080
0.000	0.049	0.080	0.200	0.080

Number of Simulators Used: 5

Remarks:
Time-Date changed.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D Lyons
DANIEL E LYONS
Signature and Printed Name

11/07/2017
Date

FDLE/ATP Form 40 – March 2004

80-000968 SMB

RECEIVED
OCT 19 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS
Serial Number: 80-000968
Time of Inspection:20:45 Date of Inspection:10/18/2017 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: 163437 Exp: 10/30/2018
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.079	0.199	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D. Lyons

DANIEL E LYONS

Signature and Printed Name

10/18/2017

Date

FDLE/ATP Form 40 -- March 2004

Reprinted From Database

80-000968 FDLE

RECEIVED
SEP 28 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS
Time of Inspection: 14:51 **Date of Inspection:** 09/28/2017
Serial Number: 80-000968
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:163437 Exp: 10/30/2018
0.000	0.000 / 0.047	0.075	0.192	0.079
0.000	0.000 / 0.046	0.075	0.194	0.079
0.000	0.000 / 0.047	0.076	0.194	0.080

Number of Simulators Used: 5

Remarks:
05: Control Outside Tolerance.

ADJUSTED FITTINGS

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D Lyons

DANIEL E LYONS
Signature and Printed Name

09/28/2017
Date

FDLE/ATP Form 40 – March 2004

80-000968 DMB

RECEIVED
AUG 29 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS
Time of Inspection: 22:11 Date of Inspection: 08/28/2017
Serial Number: 80-000968
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:163437 Exp: 10/30/2018
0.000	0.048	0.078	0.200	0.080
0.000	0.048	0.078	0.200	0.080
0.000	0.049	0.078	0.202	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D. Lyons

DANIEL E LYONS
Signature and Printed Name

08/28/2017
Date

FDLE/ATP Form 40 – March 2004

80-000968 DWS

Florida Department of Law Enforcement
Alcohol Testing Program

RECEIVED
AUG 04 2017
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS
Time of Inspection: 11:54 Date of Inspection: 08/01/2017
Serial Number: 80-000968
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:163437 Exp: 10/30/2018
0.000	0.048	0.077	0.198	0.080
0.000	0.048	0.078	0.198	0.080
0.000	0.049	0.078	0.200	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D. Lyons
DANIEL E LYONS
Signature and Printed Name

08/01/2017
Date

FDLE/ATP Form 40 – March 2004

80-000968 DWS

RECEIVED
AUG 02 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS
Time of Inspection: 11:54 Date of Inspection: 08/01/2017
Serial Number: 80-000968
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:163437 Exp: 10/30/2018
0.000	0.048	0.077	0.198	0.080
0.000	0.048	0.078	0.198	0.080
0.000	0.049	0.078	0.200	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D. Lyons

DANIEL E LYONS
Signature and Printed Name

08/01/2017
Date

FDLE/ATP Form 40 – March 2004

80-000968 DNB

RECEIVED
JUL 11 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS
Time of Inspection: 14:56 Date of Inspection: 07/11/2017
Serial Number: 80-000968
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:163437 Exp: 10/30/2018
0.000	0.048	0.078	0.199	0.081
0.000	0.048	0.079	0.200	0.080
0.000	0.049	0.079	0.201	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D. Lyons
DANIEL E LYONS
Signature and Printed Name

07/11/2017
Date

FDLE/ATP Form 40 – March 2004

80-000968 DR3

RECEIVED
JUL 11 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS
Time of Inspection: 14:04 Date of Inspection: 07/11/2017
Serial Number: 80-000968
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: 5

Remarks:
A F / M A: .

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D. Lyons
DANIEL E LYONS
Signature and Printed Name

07/11/2017
Date

FDLE/ATP Form 40 – March 2004

80-000968 RMB

RECEIVED
JUL 12 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS
Time of Inspection: 14:04 Date of Inspection: 07/11/2017
Serial Number: 80-000968
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: 5

Remarks:
A F / M A: .

DE 07/11/17 SNM/MA SOLUTION NOT USED FOR SAMPLE

The above instrument complies (~~X~~) does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D. Lyons
DANIEL E LYONS
Signature and Printed Name

07/11/2017
Date

FDLE/ATP Form 40 – March 2004

80-000968 Ems

RECEIVED
JUN 28 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS
Time of Inspection:20:44 Date of Inspection:06/27/2017
Serial Number: 80-000968
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: 163437 Exp: 10/30/2018
0.000	0.049	0.082	0.204	0.079
0.000	0.051	0.083	0.204	0.080
0.000	0.050	0.083	0.203	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DANIEL E LYONS
Signature and Printed Name

06/27/2017
Date

FDLE/ATP Form 40 -- March 2004

Reprinted From Database

80-000968 GB

RECEIVED
JUN 28 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS
Time of Inspection:20:00 Date of Inspection:06/27/2017
Serial Number: 80-000968
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: Exp:
0.000	0.050	0.079	0.189 / 0.000	
0.000	0.051	0.079	0.195 / 0.000	
0.000	0.050	0.078	0.198 / 0.000	

Number of Simulators Used: 5

Remarks:
A F / M A: No Sample Provided. 20:VALVE LOOSE,RESTART IN ERROR. Non-compliance:.20 RESTART IN ERROR.

The above instrument complies ( ) does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DANIEL E LYONS
Signature and Printed Name

06/27/2017
Date

FDLE/ATP Form 40 -- March 2004

Reprinted From Database

80-000968 DW3

RECEIVED
JUN 28 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS
Time of Inspection:18:43 Date of Inspection:06/27/2017
Serial Number: 80-000968
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: Exp:

Number of Simulators Used: _____

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DANIEL E LYONS
Signature and Printed Name

06/27/2017
Date

FDLE/ATP Form 40 -- March 2004

Reprinted From Database