

SP1006

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: INDIALANTIC P.D.
Time of Inspection: 15:20

Date of Inspection: 02/17/2017

Serial Number: 80-001006
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01316080A1 Exp: 02/05/2018
0.000	0.051	0.082	0.199	0.079
0.000	0.051	0.082	0.201	0.079
0.000	0.051	0.081	0.201	0.079


Number of Simulators Used: 3

Remarks:

Time-Date changed. TOO MUCH MOUTH ALCOHOL USED. Mouth Alcohol too much USED 1ST Attempt Repeated without incident. Check 4 mins off. Time changed

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 #9

RYAN M DOVEL

Signature and Printed Name

02/17/2017
Date