

1008
SR

RECEIVED

JAN 30 2018

Florida Department of Law Enforcement Alcohol Testing Program

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: WEST MELBOURNE P.D.
Time of Inspection: 12:33

Date of Inspection: 12/25/2017

Serial Number: 80-001008
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:04216080A3 Exp: 03/05/2018
0.000	0.049	0.077	0.193	0.079
0.000	0.049	0.078	0.194	0.079
0.000	0.049	0.078	0.193	0.079

Number of Simulators Used: 4

Remarks:

It was found that the mouth Alcohol solution used for this inspection expired on 03-27-17. This did not effect the inspection outcome. AMENDED 01-29-18. *[Signature]*

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature] #727 SHANE Z POPA
Signature and Printed Name

12/25/2017
Date

1008
SP

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JAN 30 2018

AGENCY INSPECTION REPORT - INTOXILYZER 8000 ^{FDLE} Alcohol Testing Program

Agency: WEST MELBOURNE P.D.
Time of Inspection: 10:46

Date of Inspection: 11/27/2017


Serial Number: 80-001008
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:04216080A3 Exp: 03/05/2018
0.000	0.051	0.079	0.196	0.079
0.000	0.051	0.080	0.197	0.079
0.000	0.050	0.079	0.198	0.079


Number of Simulators Used: 4

Remarks:

*It was found that the mouth alcohol solution used for this inspection expired on 03-27-17. This did not effect the inspection outcome.
AMENDED 01-29-18. *

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

OFC  #727 SHANE Z POPA
Signature and Printed Name

11/27/2017
Date

1008
SP

RECEIVED

JAN 30 2018

Florida Department of Law Enforcement Alcohol Testing Program

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: WEST MELBOURNE P.D.
Time of Inspection: 08:25

Date of Inspection: 10/17/2017

Serial Number: 80-001008
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 04216080A3 Exp: 03/05/2018
0.000	0.048	0.079	0.197	0.078
0.000	0.048	0.079	0.197	0.078
0.000	0.049	0.079	0.198	0.079

Number of Simulators Used: 3

Remarks:

It was found that the mouth Alcohol solution used for this inspection expired on 03-27-17. This did not effect the inspection outcome. AMENDED 01-29-18. [Signature]

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature] SHANE Z POPA
Signature and Printed Name

10/17/2017
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WEST MELBOURNE P.D.
Time of Inspection: 12:36

Date of Inspection: 10/05/2017

Serial Number: 80-001008
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

RMS

N/A Compliance Not Determined

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

10/05/2017
Date

10/10/17 JD

1008 SP

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

30 2018

AGENCY INSPECTION REPORT - INTOXILYZER 8000 FDLE

Agency: WEST MELBOURNE P.D.
Time of Inspection: 20:55

Date of Inspection: 08/30/2017

Serial Number: 80-001098
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:04216080A3 Exp: 03/05/2018
0.000	0.049	0.080	0.199	0.078
0.000	0.049	0.080	0.201	0.078
0.000	0.049	0.081	0.201	0.078

Number of Simulators Used: 3

Remarks:

It was found that the mouth Alcohol solution used for this inspection expired on 03-27-17. This did not effect the inspection outcome.
AMENDED 01-29-18 *AS*

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shane Z Popa #727 SHANE Z POPA
Signature and Printed Name

08/30/2017
Date

1008
SP

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JAN 30 2018

AGENCY INSPECTION REPORT - INTOXILYZER 8000 FDLE Alcohol Testing Program

Agency: WEST MELBOURNE P.D.
Time of Inspection: 08:44

Date of Inspection: 07/31/2017

Serial Number: 80-001008
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:04216080A3 Exp: 03/05/2018
0.000	0.049	0.081	0.189 / 0.199	0.083
0.000	0.050	0.081	0.195 / 0.200	0.082
0.000	0.050	0.081	0.198 / 0.200	0.082

Number of Simulators Used: 3

Remarks:

20: Control Outside Tolerance.

- During the 0.20g/210L test, the first sample was 0.189 and all others were within range. The 0.20g/210L test was repeated without any issues. The inspection was completed with no further issues and was in compliance with Chapter 11D-8 FAC. Possible cause was not enough time for solution to warm up. Corrective action was to allow solution to warm up more before the test was repeated.
- It was found that the mouth alcohol solution used for this inspection expired on 03-27-17. This did not effect the outcome for this inspection. AMENDED 01-29-18

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature]

Signature and Printed Name

SHANE Z POPA

07/31/2017
Date

1008
SP

RECEIVED

JAN 30 2018

Florida Department of Law Enforcement Alcohol Testing Program

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: WEST MELBOURNE P.D.
Time of Inspection: 09:40

Date of Inspection: 06/25/2017

Serial Number: 80-001008
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:04216080A3 Exp: 03/05/2018
0.000	0.049	0.081	0.197	0.079
0.000	0.050	0.080	0.199	0.079
0.000	0.050	0.081	0.200	0.079

Number of Simulators Used: 3

Remarks:

It was found that the mouth Alcohol solution used for this inspection expired on 03-27-17. This did not effect the inspection outcome. AMENDED 01-29-18 *B*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shane Z Popa #727

SHANE Z POPA

Signature and Printed Name

06/25/2017
Date

1008
SR

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JAN 30 2018

AGENCY INSPECTION REPORT - INTOXILYZER 8000 FDLE Alcohol Testing Program

Agency: WEST MELBOURNE P.D.
Time of Inspection: 12:53

Date of Inspection: 05/24/2017

Serial Number: 80-001008
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A1 Exp: 10/05/2017
0.000	0.048	0.079	0.197	0.080
0.000	0.049	0.080	0.198	0.079
0.000	0.050	0.080	0.198	0.080

Number of Simulators Used: 3

Remarks:

It was found that the mouth Alcohol solution used for this inspection expired on 03-27-17. This did not effect the inspection outcome. AMENDED 01-29-18. *AS*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shane Z Popa

SHANE Z POPA

Signature and Printed Name

05/24/2017
Date

1008
SP

RECEIVED

JAN 30 2018

Florida Department of Law Enforcement Alcohol Testing Program

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WEST MELBOURNE P.D.
Time of Inspection: 15:27

Date of Inspection: 04/30/2017

Serial Number: 80-001008
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A1 Exp: 10/05/2017
0.000	0.049	0.080	0.199	0.080
0.000	0.050	0.081	0.201	0.079
0.000	0.049	0.081	0.200	0.079

Number of Simulators Used: 3

Remarks:

It was found that the mouth Alcohol solution that was used for this inspection expired on 03-27-17. This did not effect the inspection outcome. AMENDED 01-29-18. *[Signature]*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature]

SHANE Z POPA

Signature and Printed Name

04/30/2017
Date

Florida Department of Law Enforcement

Alcohol Testing Program

JAN 3 0 2018
RECEIVED

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WEST MELBOURNE P.D. Serial Number: 80-001008
 Time of Inspection: 08:23 Date of Inspection: 03/28/2017 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK	Yes	No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 24515080A1 Exp: 10/05/2017
0.000	0.049	0.080	0.189 / 0.197	0.080
0.000	0.049	0.080	0.194 / 0.198	0.079
0.000	0.050	0.081	0.196 / 0.199	0.079

Number of Simulators Used: 3
 Remarks:
 20: Control Outside Tolerance.

01-29-18/It was found that the mouth Alcohol Solution that was used for this inspection expired on 03-27-17. This did not effect the inspection. AMENDED 01-29-18

During the 0.20g/210L Test the first sample was 0.189, all others were within range. The 0.20g/210L Test was repeated without any issues. The inspection was completed with no further issues and in compliance with 11D-8. Possible cause was not enough time for solution to warm up. Corrective action was to allow second solution to warm up for 30 minutes.

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC. AMENDED 05-24-17

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shane Z Popa #727 SHANE Z POPA
 Signature and Printed Name

03/28/2017
 Date



Florida Department of Law Enforcement

AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: WEST MELBOURNE PD Instrument Serial Number: 80-001008

AGENCY INSPECTION DATA REVIEW		
Agency Inspector: <u>SHANE POPA</u>	Date of Inspection: <u>3/28/17</u>	Time of Inspection: <u>8:23</u>
Agency Inspection Discrepancy: <input checked="" type="checkbox"/> Incomplete <input type="checkbox"/> Untimely/Not Received <input type="checkbox"/> Erroneous Information <input type="checkbox"/> Procedural <input type="checkbox"/> Other _____		
<input type="checkbox"/> Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.		
<input type="checkbox"/> Lot Number <input type="checkbox"/> Expiration Date for _____ g/210L <input type="checkbox"/> Alcohol Reference Solution <input type="checkbox"/> Dry Gas Standard is <input type="checkbox"/> Incorrect <input type="checkbox"/> Expired.		
<input checked="" type="checkbox"/> FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report – Intoxilyzer 8000. The <input type="checkbox"/> REASON for repeating the following test(s); OR the <input checked="" type="checkbox"/> Possible Cause and Corrective Action Taken on the following test(s) was not recorded: <input type="checkbox"/> Alcohol Free Subject Test <input type="checkbox"/> Mouth Alcohol Test <input type="checkbox"/> Alcohol Free Test <input type="checkbox"/> Interferent Detect Test <input type="checkbox"/> 0.05 g/210L Test <input type="checkbox"/> 0.08 g/210L Test <input checked="" type="checkbox"/> 0.20 g/210L Test <input type="checkbox"/> 0.08 g/210L Dry Gas Standard Test		
<input type="checkbox"/> FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector." <input type="checkbox"/> The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC. <input type="checkbox"/> The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use. <input type="checkbox"/> The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.		
<input type="checkbox"/> The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.		
<input type="checkbox"/> Other: _____		

OTHER ELECTRONIC DATA REVIEW	
<input type="checkbox"/> Login Records Date: _____	Comments: _____
<input type="checkbox"/> Cylinder Change Records Date: _____	
<input type="checkbox"/> Control Test Records Date: _____	
<input type="checkbox"/> Diagnostic Check Records Date: _____	

CORRECTIVE ACTION
<input checked="" type="checkbox"/> Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by <u>left</u> (Date).
<input type="checkbox"/> Provide a written explanation regarding the referenced item(s) to the Department Inspector by _____ (Date).
<input type="checkbox"/> Upload the Agency Inspection(s).
<input type="checkbox"/> Remove the instrument from evidentiary use until otherwise directed by the Department.
<input type="checkbox"/> No action required
<input type="checkbox"/> Other: _____

Shane Platt
Signature of Alcohol Testing Program Staff Member

5/23/2017
Date

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

Agency WEST MELBOURNE P.D.	Date 03/28/2017	Type Agency Inspection	
Inspector POPA, SHANE, Z	Time 08:23:51	Software 8100.27	
0.05 Lot # 201603D	0.08 Lot # 201611B	0.20 Lot # 201604C	0.08 Gas Lot # 24515080A1
Expiration 03/08/2018	Expiration 11/15/2018	Expiration 04/05/2018	Expiration 10/05/2017

Number of Simulators 3

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.189
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.196
		0.20 g/210L Test # (Repeat) 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:
 20: Control Outside Tolerance.

Data Download Date/Time 03/28/2017 08:31

1008 SP

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WEST MELBOURNE P.D.

Time of Inspection: 08:23

Date of Inspection: 03/28/2017

Serial Number: 80-001008

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A1 Exp: 10/05/2017
0.000	0.049	0.080	0.189 / 0.197	0.080
0.000	0.049	0.080	0.194 / 0.198	0.079
0.000	0.050	0.081	0.196 / 0.199	0.079

Number of Simulators Used: 3

Remarks:

20: Control Outside Tolerance.

During the 0.20g/210L Test the first sample was 0.189, all others were within range. The 0.20g/210L Test was repeated without any issues. The inspection was completed with no further issues and in compliance with 11D-8. Possible cause was not enough time for solution to warm up. Corrective action was to allow second solution to warm up for 30 minutes.

The above instrument complies () does not comply () with Chapter 11D-8, FAC. AMENDED 05-24-17

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature]

SHANE Z POPA

Signature and Printed Name

03/28/2017

Date