

1018 RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

JUN 16 2017

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: JACKSONVILLE SO
Time of Inspection: 10:20

Date of Inspection: 06/15/2017

Serial Number: 80-001018
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:17316080A3 Exp: 08/05/2018
0.000	/ 0.046	0.072 / 0.075	0.198	0.078
0.000	/ 0.046	0.073 / 0.075	0.196	0.078
0.000	/ 0.046	0.073 / 0.076	0.196	0.078

Number of Simulators Used: 5

Remarks:
05: RFI Detect. 08:NO RFI SOURCE IN RM RADIO USE IN HALL

AMMENDMENT: *MP 1599*
PLACED INTO SERVICE/ON LOAN TO NAS MAYPORT 06/15/2017

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MP 1599

MICHAEL R POTTER
Signature and Printed Name

06/15/2017
Date

1018 (3)

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: JACKSONVILLE SO
Time of Inspection: 10:12

Date of Inspection: 02/24/2017

Serial Number: 80-001018
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.048	0.077	0.180 / 0.189	
0.000	0.047	0.076	0.185 / 0.191	
0.000	0.048	0.077	0.188 / 0.192	

Number of Simulators Used: 5

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:02 TOL/REMOVED FROM SER.

- AI called me to inform me of failed inspection.
He will take out of service + send to repair.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature]
MP-7599

Signature and Printed Name

MICHAEL R POTTER

02/24/2017
Date