

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**

AUG 08 2017

FDLE

## AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: COCONUT CREEK PD  
Time of Inspection: 12:30

Date of Inspection: 07/26/2017

Serial Number: 80-001044  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 686331 Exp: 01/31/2020
0.000	0.049	0.078	0.197	0.080
0.000	0.050	0.078	0.197	0.080
0.000	0.050	0.077	0.197	0.080

Number of Simulators Used: 4

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

KYLE DUFFEY

Signature and Printed Name

07/26/2017  
Date

RECEIVED

JUL 05 2017

# Florida Department of Law Enforcement

## Alcohol Testing Program

FDLE  
Alcohol Testing Program

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COCONUT CREEK PD  
Time of Inspection: 11:01

Date of Inspection: 06/26/2017

Serial Number: 80-001044  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:686331 Exp: 01/31/2020
0.000	0.047	0.077	0.196	0.080
0.000	0.047	0.077	0.196	0.080
0.000	0.048	0.077	0.196	0.080

Number of Simulators Used: 4

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
Signature and Printed Name

KYLE DUFFEY

06/26/2017  
Date

1044  
B

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COCONUT CREEK PD  
Time of Inspection: 12:56

Date of Inspection: 04/24/2017

Serial Number: 80-001044  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1694353 Exp: 06/10/2017
0.000	0.047	0.076	0.193	0.080
0.000	0.047	0.077	0.196	0.080
0.000	0.048	0.078	0.197	0.080

Number of Simulators Used: 4

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

KYLE DUFFEY

Signature and Printed Name

04/24/2017  
Date

80-001044 DBB

# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

MAR 31 2017

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COCONUT CREEK PD  
Time of Inspection: 11:11

Date of Inspection: 03/21/2017

Serial Number: 80-001044  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1694353 Exp: 06/10/2017
0.000	0.047	0.076	0.194	0.081
0.000	0.048	0.076	0.194	0.080
0.000	0.047	0.076	0.195	0.082

Number of Simulators Used: 4

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

KYLE DUFFEY

Signature and Printed Name

03/21/2017  
Date

1044

# Florida Department of Law Enforcement

## Alcohol Testing Program

RECEIVED

MAR 03 2017

FDLE

Alcohol Testing Program

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COCONUT CREEK PD  
Time of Inspection: 11:24

Date of Inspection: 02/21/2017

Serial Number: 80-001044  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1694353 Exp: 06/10/2017
0.000	0.047	0.077	0.196	0.080
0.000	0.048	0.076	0.196	0.080
0.000	0.048	0.077	0.195	0.081

Number of Simulators Used: 4

Remarks:

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

KYLE DUFFEY

Signature and Printed Name

02/21/2017  
Date

1044  
R

RECEIVED

FEB 28 2017

# Florida Department of Law Enforcement

## Alcohol Testing Program

FDLE  
Alcohol Testing Program

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COCONUT CREEK PD  
Time of Inspection: 10:38

Date of Inspection: 02/13/2017

Serial Number: 80-001044  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1694353 Exp: 06/10/2017
0.000	0.048	0.074 / 0.075	0.196	0.079
0.000	0.049	0.075 / 0.076	0.196	0.079
0.000	0.048	0.076 / 0.076	0.195	0.079

Number of Simulators Used: 4

Remarks:

08: Control Outside Tolerance. POSS AIR LEAK AT INTAKE TUBE



The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Kyle Duffey*  
#17979

Signature and Printed Name

KYLE DUFFEY

02/13/2017  
Date

1044  
B

# Florida Department of Law Enforcement

## Alcohol Testing Program

RECEIVED

JAN 30 2017

FDLE

Alcohol Testing Program

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COCONUT CREEK PD  
Time of Inspection: 11:02

Date of Inspection: 01/24/2017

Serial Number: 80-001044  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1694353 Exp: 06/10/2017
0.000	0.048	0.075	0.195	0.080
0.000	0.048	0.076	0.194	0.080
0.000	0.049	0.076	0.195	0.080

Number of Simulators Used: 4

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 #17979  
\_\_\_\_\_  
Signature and Printed Name

KYLE DUFFEY

01/24/2017  
\_\_\_\_\_  
Date