Florida Department of Law Enforcement
Alcohol Testing Program

1054
SP

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 08:10
Date of Inspection: 12/30/2017
Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32316080A3 Exp: 01/05/2019
0.000	0.050	0.078	0.200	0.080
0.000	0.048	0.078	0.199	0.079
0.000	0.049	0.078	0.199	0.079

Number of Simulators Used: 5

Remarks:

A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY Shon W. Abeln 2184 D.S.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184
SHON W ABELN
Signature and Printed Name

12/30/2017
Date

FDLE/ATP Form 40 – March 2004

1054
SP

RECEIVED
DEC 04 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 08:05
Date of Inspection: 11/28/2017
Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32316080A3 Exp: 01/05/2019
0.000	0.048	0.077	0.197	0.079
0.000	0.048	0.078	0.198	0.079
0.000	0.047	0.078	0.197	0.079

Number of Simulators Used: 5

Remarks:

A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY Shon W. Abeln 2184 D.S.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184
SHON W ABELN
Signature and Printed Name

11/28/2017
Date

FDLE/ATP Form 40 – March 2004

1054 SP

RECEIVED
OCT 25 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 10:18
Date of Inspection: 10/21/2017
Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32316080H3 Exp: 01/15/2019
0.000	0.047	0.078	0.198	0.079
0.000	0.048	0.078	0.198	0.078
0.000	0.048	0.078	0.198	0.079

Number of Simulators Used: 5

Remarks:

A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY 2184 D.S.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

2184

SHON W ABELN
Signature and Printed Name

10/21/2017
Date

FDLE/ATP Form 40 – March 2004

1054 SP

RECEIVED
SEP 25 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 14:33
Date of Inspection: 09/20/2017
Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32316080A3 Exp: 01/05/2019
0.000	0.046	0.076	0.194	0.078
0.000	0.047	0.076	0.197	0.078
0.000	0.048	0.077	0.197	0.078

Number of Simulators Used: 5

Remarks:

A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY_____D.S.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GREGORY S GLASCO
Signature and Printed Name

09/20/2017
Date

FDLE/ATP Form 40 – March 2004

Florida Department of Law Enforcement
Alcohol Testing Program

1054 SP

RECEIVED
AUG 29 2017
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 12:53
Date of Inspection: 08/24/2017
Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32316080A3 Exp: 01/05/2019
0.000	0.047	0.078	0.192	0.078
0.000	0.047	0.078	0.193	0.079
0.000	0.048	0.077	0.194	0.078

Number of Simulators Used: 5

Remarks:

A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY_____D.S.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GREGORY S GLASCO
Signature and Printed Name

08/24/2017
Date

FDLE/ATP Form 40 – March 2004

RECEIVED
AUG 02 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 08:20
Date of Inspection: 08/02/2017
Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:
BYPASSED AI TO OPERATE INSTRUMENT

PJM

N/A COMPLIANCE
NOT DETERMINED SP

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt
SHAYLA D PLATT
Signature and Printed Name

08/02/2017
Date

JD

1054
SP

RECEIVED
JUN 26 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 14:48
Date of Inspection: 06/18/2017
Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A3 Exp: 10/05/2017
0.000	0.048	0.079	0.197	0.078
0.000	0.048	0.079	0.195	0.078
0.000	0.048	0.079	0.195	0.078

Number of Simulators Used: 5

Remarks:

A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY Shon W. Abeln 2184 D.S.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184
SHON W ABELN
Signature and Printed Name

06/18/2017
Date

FDLE/ATP Form 40 – March 2004

1054
SP

RECEIVED
JUN 14 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 08:01
Date of Inspection: 05/28/2017
Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A3 Exp: 10/05/2017
0.000	0.050	0.079	0.200	0.079
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.079	0.199	0.078

Number of Simulators Used: 5

Remarks:
A F / M A: RFI Detect.

My tail Radio was ON and when I Realized it I Turned it off and got The R.F.I.

Sue
Amended 6/9/2017 @ 9:37 AM

A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY Shon W. Abeln 2184 D.S.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184
SHON W ABELN
Signature and Printed Name

05/28/2017
Date

FDLE/ATP Form 40 – March 2004

FDLE
Florida Department of Law Enforcement

AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Highlands County SO

Instrument Serial Number: 80-001054

AGENCY INSPECTION DATA REVIEW

Agency Inspector: Shon Abeln	Date of Inspection: 5/28/17	Time of Inspection: 8:01

Agency Inspection Discrepancy: ☒ Incomplete ☐ Untimely/Not Received ☐ Erroneous Information ☐ Procedural ☐ Other _____

☐ Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.

☐ Lot Number ☐Expiration Date for _____g/210L ☐Alcohol Reference Solution ☐Dry Gas Standard is ☐Incorrect ☐Expired.

☒ FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report – Intoxilyzer 8000. The ☐REASON for repeating the following test(s); OR the ☒ Possible Cause and Corrective Action Taken on the following test(s) was not recorded:

☐ Alcohol Free Subject Test ☒ Mouth Alcohol Test ☐ Alcohol Free Test ☐ Interferent Detect Test
☐ 0.05 g/210L Test ☐ 0.08 g/210L Test ☐ 0.20 g/210L Test ☐ 0.08 g/210L Dry Gas Standard Test

☐ FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector."

- ☐ The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC.
- ☐ The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use.
- ☐ The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.

☐ The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.

☐ Other: _____

OTHER ELECTRONIC DATA REVIEW

	Comments:
☐ Login Records Date: ____	_____
☐ Cylinder Change Records Date: ____	
☐ Control Test Records Date: ____	
☐ Diagnostic Check Records Date: ____	

CORRECTIVE ACTION

☒ Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by 6/22/17 (Date).
☐ Provide a written explanation regarding the referenced item(s) to the Department Inspector by _____ (Date).
☐ Upload the Agency Inspection(s).
☐ Remove the instrument from evidentiary use until otherwise directed by the Department.
☐ No action required
☐ Other: _____

Shayla Platt
Signature of Alcohol Testing Program Staff Member

6/8/2017
Date

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

Agency HIGHLANDS COUNTY SO **Date** 05/28/2017 **Type** Agency Inspection
Inspector ABELN,SHON,W **Time** 08:01:50 **Software** 8100.27

0.05 Lot # 201603D **0.08 Lot #** 201611B **0.20 Lot #** 201604C **0.08 Gas Lot #** 24515080A3
Expiration 03/08/2018 **Expiration** 11/15/2018 **Expiration** 04/05/2018 **Expiration** 10/05/2017

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: RFI Detect.

Data Download Date/Time 05/28/2017 08:08

Printed: 08 June 2017

COBRA ©1997-2014 CMI, Inc.

1054 SR

RECEIVED
JUN 01 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 08:01
Date of Inspection: 05/28/2017
Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A3 Exp: 10/05/2017
0.000	0.050	0.079	0.200	0.079
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.079	0.199	0.078

Number of Simulators Used: 5

Remarks:
A F / M A: RFI Detect.

A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY Shon W. Abeln 2184 D.S.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184 SHON W ABELN
Signature and Printed Name

05/28/2017
Date

FDLE/ATP Form 40 – March 2004

1054 SP

RECEIVED
MAY 04 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 11:45
Date of Inspection: 04/23/2017
Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A3 Exp: 10/05/2017
0.000	0.048	0.078	0.199	0.078
0.000	0.048	0.078	0.199	0.078
0.000	0.048	0.077	0.197	0.077

Number of Simulators Used: 5

Remarks:

A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY Shon W. Abeln 2184 D.S.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184 SHON W ABELN
Signature and Printed Name

04/23/2017
Date

FDLE/ATP Form 40 – March 2004

1054 SP

RECEIVED
MAR 23 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 15:16
Date of Inspection: 03/20/2017
Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A3 Exp: 10/05/2017
0.000	0.048	0.077	0.192	0.080
0.000	0.048	0.078	0.193	0.079
0.000	0.047	0.078	0.194	0.079

Number of Simulators Used: 5

Remarks:

A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY_____D.S.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GREGORY S GLASCO
Signature and Printed Name

03/20/2017
Date

FDLE/ATP Form 40 – March 2004

1054
8P

RECEIVED
FEB 22 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 13:28
Date of Inspection: 02/17/2017
Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A3 Exp: 10/05/2017
0.000	0.048	0.078	0.194	0.079
0.000	0.049	0.077	0.195	0.079
0.000	0.049	0.077	0.196	0.079

Number of Simulators Used: 5

Remarks:

A TRUE COPY
PAUL BLACKMAN, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY_____D.S.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GREGORY S GLASCO
Signature and Printed Name

02/17/2017
Date

FDLE/ATP Form 40 – March 2004

1054
SP

RECEIVED
JAN 27 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 14:16
Date of Inspection: 01/25/2017
Serial Number: 80-001054
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A3 Exp: 10/05/2017
0.000	0.048	0.077	0.195	0.078
0.000	0.049	0.077	0.195	0.079
0.000	0.048	0.078	0.196	0.078

Number of Simulators Used: 5

Remarks:

P. Blackman

A TRUE COPY
SUSAN BENTON, SHERIFF
HIGHLANDS COUNTY, FLORIDA
BY _____ D.S.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GREGORY S GLASCO
Signature and Printed Name

01/25/2017
Date

FDLE/ATP Form 40 – March 2004