

* UPDATED 02/29/2018 *W2674*

80-001068 *W2674*

RECEIVED

MAR 01 2018

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLEARWATER PD
Time of Inspection: 20:40

Date of Inspection: 12/21/2017

Serial Number: 80-001068
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) <i>201604C</i> Lot#:201611B Exp: 11/15/2018 <i>04/05/2018</i>	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:26316080A2 Exp: 11/05/2018
0.000	0.047	0.078	0.193 <i>0124/17</i>	0.079
0.000	0.049	0.079	0.196	0.079
0.000	0.049	0.079	0.198	0.080

Number of Simulators Used: 5

Remarks:
CHANGE TIME

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DANIEL M NEGERSMITH

Signature and Printed Name

12/21/2017
Date

80-001068 *DNB*

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

JUN 01 2017

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLEARWATER PD
Time of Inspection: 15:23

Date of Inspection: 05/26/2017

Serial Number: 80-001068
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:19615080A2 Exp: 09/05/2017
0.000	0.048	0.079	0.194	0.079
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.080	0.197	0.079


Number of Simulators Used: 5

Remarks:

ALCOHOL LOT 2016C ACETONE LOT 2016B

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

DAVID A NUGENT

05/26/2017
Date