

1077  
SP

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HARDEE COUNTY SO  
Time of Inspection: 11:11

Date of Inspection: 07/26/2017

Serial Number: 80-001077  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 203498 Exp: 12/10/2018
0.000	0.049	0.080	0.197	0.079
0.000	0.049	0.080	0.197	0.079
0.000	0.050	0.080	0.198	0.079

Number of Simulators Used: 5

Remarks:  
A F / M A: Mouth alcohol residue in mouth on first two test.  
Instrument sent out for repairs

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 \_\_\_\_\_  
Signature and Printed Name

GARY W. CARTWRIGHT JR.

07/26/2017  
Date