

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
APR 28 2017
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 21:37

Date of Inspection: 04/28/2017

Serial Number: 80-001109
Software: 8100.27

| Check or Test | YES | NO |
|---------------------------------------------|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |


| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/03/2018 | 0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018 | 0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 09216080A3 Exp: 05/05/2018 |
|----------------------------|--------------------------------------------------------------|--------------------------------------------------------------|--------------------------------------------------------------|------------------------------------------------------------------------------|
| 0.000 | 0.047 | 0.077 | 0.193 | 0.079 |
| 0.200 | 0.048 | 0.078 | 0.195 | 0.078 |
| 0.000 | 0.048 | 0.079 | 0.195 | 0.078 |

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JOSHUA C CAMPBELL

Signature and Printed Name

04/28/2017
Date

80-001109 008

RECEIVED

MAR 20 2017

FDLE Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 21:09

Date of Inspection: 03/20/2017

Serial Number: 80-001109
Software: 8100.27

| Check or Test | YES | NO |
|---------------------------------------------|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |


| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018 | 0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018 | 0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:09216080A3 Exp: 05/05/2018 |
|----------------------------|-------------------------------------------------------------|-------------------------------------------------------------|-------------------------------------------------------------|-----------------------------------------------------------------------------|
| 0.000 | 0.050 | 0.080 | 0.194 | 0.080 |
| 0.000 | 0.050 | 0.081 | 0.197 | 0.080 |
| 0.000 | 0.050 | 0.080 | 0.198 | 0.079 |

Number of Simulators Used: 5

Remarks:
NOT ENOUGH SOLUTION.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 Signature and Printed Name JOSHUA C. CAMPBELL / J.C.A. Hernandez

03/20/2017
Date

80-001109 *KMS*

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
FEB 28 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 23:04

Date of Inspection: 02/27/2017

Serial Number: 80-001109
Software: 8100.27

| Check or Test | YES | NO |
|---------------------------------------------|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017 | 0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018 | 0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 09216080A3 Exp: 05/05/2018 |
|----------------------------|--------------------------------------------------------------|--------------------------------------------------------------|--------------------------------------------------------------|------------------------------------------------------------------------------|
| 0.000 | 0.050 | 0.079 | 0.196 | 0.080 |
| 0.000 | 0.050 | 0.079 | 0.198 | 0.079 |
| 0.000 | 0.051 | 0.080 | 0.198 | 0.079 |

Number of Simulators Used: 5

Remarks:
TOO MUCH M/A.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JOSHUA C CAMPBELL
Signature and Printed Name

02/27/2017
Date

80-001109 *RMS*

RECEIVED

JAN 25 2017

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 22:41

Date of Inspection: 01/25/2017

Serial Number: 80-001109
Software: 8100.27

| Check or Test | YES | NO |
|---------------------------------------------|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017 | 0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018 | 0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 09216080A3 Exp: 05/05/2018 |
|----------------------------|--------------------------------------------------------------|--------------------------------------------------------------|--------------------------------------------------------------|------------------------------------------------------------------------------|
| 0.000 | 0.049 | 0.079 | 0.197 | 0.080 |
| 0.000 | 0.050 | 0.080 | 0.198 | 0.079 |
| 0.000 | 0.050 | 0.080 | 0.198 | 0.079 |

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

JOSHUA C. CAMPEELI

01/25/2017
Date