

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**

AUG 21 2017

**FDLE**

## AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: PONCE INLET PD  
Time of Inspection: 08:43

Date of Inspection: 08/18/2017

Serial Number: 80-001158  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_

**Remarks:**

SKIPPED AI TO OPERATE INSTRUMENT

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

PATRICK J MURPHY

\_\_\_\_\_  
Signature and Printed Name

08/18/2017  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**

APR 11 2017

**FDLE**

Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Ponce Inlet PD  
Time of Inspection: 0724

Date of Inspection: 03/31/2017

Serial Number: 80-001158  
Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
N/A	N/A	N/A	N/A	N/A
↓	↓	↓	↓	↓

Number of Simulators Used: 0

Remarks: UNABLE TO POWER ON,  
UNABLE TO COMPLETE INSPECTION PRIOR TO SENDING TO  
FDLE  
TAKEN OUT OF SERVICE SENT TO FDLE

The above instrument complies (      ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

CPL AG PI 1314

William C. Ryan

Signature and Printed Name

03/31/2017  
Date