

80-001174 EAS

RECEIVED

MAR 30 2017

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLERMONT P.D.

Serial Number: 80-001174

Time of Inspection: 11:41

Date of Inspection: 03/30/2017

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:686331 Exp: 01/31/2020
0.000	0.051	0.079	0.196	0.080
0.000	0.052	0.080	0.197	0.081
0.000	0.052	0.080	0.198	0.081

Number of Simulators Used: 4

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DEBBIE REASONER

Signature and Printed Name

03/30/2017
Date

RDS 80-001174

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

FEB 07 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLERMONT P.D.
Time of Inspection: 15:24

Date of Inspection: 02/07/2017

Serial Number: 80-001174
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:477627 Exp: 06/29/2019
0.000	0.051	0.079	0.196	0.080
0.000	0.050	0.080	0.197	0.081
0.000	0.051	0.080	0.198	0.081

Number of Simulators Used: 4

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 191
 DEBBIE REASONER
 Signature and Printed Name

02/07/2017
Date

80-001174

RECEIVED
JAN 25 2017

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLERMONT P.D. Date of Inspection: 01/25/2017 Serial Number: 80-001174
Time of Inspection: 10:34 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK	Yes	No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:477627 Exp: 06/29/2019
0.000	0.050	0.080	0.197	0.080
0.000	0.051	0.080	0.198	0.080
0.000	0.052	0.080	0.198	0.080

Number of Simulators Used: 4

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 141
Signature and Printed Name DEBBIE REASONER

01/25/2017
Date