

Florida Department of Law Enforcement Alcohol Testing Program

11898

RECEIVED

APR 20 2017

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: MARTIN COUNTY SO
Time of Inspection: 13:22

Date of Inspection: 03/28/2017

Serial Number: 80-001189
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:06616080A2 Exp: 04/05/2018
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.081	0.202	0.079
0.000	0.050	0.081	0.202	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name BRADFORD B BAKER

03/28/2017
 Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

APR 2

FI

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Test

Serial Number: 80-00
Software: 8100.27

Agency: MARTIN COUNTY SO
Time of Inspection: 11:40

Date of Inspection: 03/28/2017

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std (g/210L) Lot#: Exp:

Number of Simulators Used: 5

Remarks:
A F / M A: . Non-compliance: .

M/A TEST 0.192, too little m
Repeat Inspection

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permissions of Chapter 11D-8, FAC.

BRADFORD B BAKER

Signature: _____ and Printed Name

03/28/2017
Date