

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

MAR 21 2017

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO  
Time of Inspection: 17:29

Date of Inspection: 03/20/2017

Serial Number: 80-001249  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:34015080A3 Exp: 01/05/2018
0.000	0.050	0.078	0.191	0.086 / 0.087
0.000	0.050	0.079	0.196	0.086 / 0.086
0.000	0.050	0.079	0.197	0.086 / 0.086

Number of Simulators Used: 5

Remarks:

08: Control Outside CHANGED GAS TANK AND O RINGS., Control . WASHINGTON AND FRAZIER OBSERVED.

*Changed gas: Lot# 562838; Exp. 09/06/2019*

The above instrument complies ( ~~X~~ ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Franca R. Dicienzo*

FRANCA R DICIZENZO

Signature and Printed Name

03/20/2017

Date