

RECEIVED

NOV 20 2017

Florida Department of Law Enforcement Alcohol Testing Program



FDLE

Alcohol Testing Program AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: JACKSONVILLE SO
Time of Inspection: 13:40

Date of Inspection: 11/17/2017

Serial Number: 80-001279
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:26316080A2 Exp: 11/05/2018
0.000	0.049	0.080 / 0.080	0.201	0.080
0.000	0.049	0.080 / 0.080	0.202	0.080
0.000	0.050	/ 0.080	0.202	0.080

Number of Simulators Used: 5

Remarks:

08: RFI Detect.RADIO TRAFIC IN HALL. RERAN OK

AMENDMENT:

During the third run of the .08 wet-bath Simulator I received a RFI due to radio traffic in the hallway in the intake area. After verifying the source of the RFI, and clearing the area, the .08 wet-bath Simulator was re-run without incident.

Sgt. M Potter 7599

Handwritten signature and date: Sgt M Potter 7599 11.17.17

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Handwritten signature: Sgt M Potter 7599 MICHAEL R POTTER
Signature and Printed Name

11/17/2017
Date

1279 @ RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

JUN 16 2017

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: JACKSONVILLE SO
Time of Inspection: 11:18

Date of Inspection: 06/15/2017

Serial Number: 80-001279
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:26316080A2 Exp: 11/05/2018
0.000	0.051	0.078	0.197 / 0.201	0.078
0.000	0.051	0.079	RFI / 0.199	0.078
0.000	0.050	0.080	/ 0.198	0.078

Number of Simulators Used: 5

Remarks:

20: RFI Detect.NO RFI SOURCE IN RM RADIO USE IN HALL

AMMENDENT:

Placed into sevice 6/15/2017

Handwritten: 1279

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Handwritten Signature: [Signature] 1279

MICHAEL R POTTER

Signature and Printed Name

06/15/2017
Date

1279

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

APR 27 2017
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: JACKSONVILLE SO
Time of Inspection: 13:57

Date of Inspection: 04/27/2017

Serial Number: 80-001279
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01316080A2 Exp: 02/05/2018
0.000	0.051	0.080	0.205	0.082
0.000	0.051	0.080	0.203	0.082
0.000	0.050	0.078	0.202	0.082

Number of Simulators Used: 5

Remarks: REMOVED FROM SERVICE FOR ANNUAL PREVENTIVE MAINTENANCE. AMENDED

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MICHAEL R POTTER

Signature and Printed Name

04/27/2017
Date