

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

MAR 13 2017

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGH SPRINGS PD
Time of Inspection: 16:23

Date of Inspection: 02/28/2017

Serial Number: 80-001283
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.046	0.078	0.188 / 0.178	
0.000	0.047	0.078	0.000 / 0.183	
0.000	0.046	0.078	0.000 / 0.185	

Issue every time

Number of Simulators Used: 5

Remarks:

20: Control Outside BELOW RANGE/SIMULATOR FIXED RETRYING . BELOW RANGE SIMULATOR IS BEING SWITCHED OUT Non-compliance: .20 UNDER RANGE AND SIMULATOR IS A NEW ONE .

1

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAMES R YAKUBSIN

Signature and Printed Name

02/28/2017
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

APR 12 2017

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Testing Program

Agency: HIGH SPRINGS PD
Time of Inspection: 08:36

Date of Inspection: 04/12/2017

Serial Number: 80-001283
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:
BYPASS TO OPERATE INSTRUMENT

SP JB

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAKE L SHANAHAN

Signature and Printed Name

04/12/2017
Date