

1278

RECEIVED

Florida Department of Law Enforcement MAR 06 2017 Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 09:42

Date of Inspection: 02/28/2017

Serial Number: 80-001286
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32316080A3 Exp: 01/05/2019
0.000	0.047	0.078	0.199	0.081
0.000	0.048	0.078	0.200	0.081
0.000	0.048	0.078	0.199	0.081

Number of Simulators Used: 5

Remarks:

BAKER COUNTY
SHERIFF'S OFFICE
MAR 01 2017
RECORDS SECTION
RECEIVED

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name COREY S WILSON

02/28/2017
Date

1287

RECEIVED

MAR 06 2017

Florida Department of Law Enforcement Alcohol Testing Program

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 09:36

Date of Inspection: 02/28/2017

Serial Number: 80-001287
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32316080A3 Exp: 01/05/2019
0.000	0.047	0.078	0.196	0.081
0.000	0.048	0.078	0.197	0.081
0.000	0.048	0.078	0.197	0.081

Number of Simulators Used: 5

Remarks:

BAKER COUNTY
SHERIFF'S OFFICE
MAR 01 2017
RECORDS SECTION
RECEIVED

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Corey S Wilson
Signature and Printed Name COREY S WILSON

02/28/2017
Date

128e @

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

FEB 06 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 20:29

Date of Inspection: 01/31/2017

Serial Number: 80-001286
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A1 Exp: 10/05/2017
0.000	0.047	0.078	0.198	0.080
0.000	0.048	0.078	0.200	0.080
0.000	0.049	0.079	0.201	0.080

Number of Simulators Used: 5

Remarks:

BAKER COUNTY
SHERIFFS OFFICE
FEB 01 2017
RECORDS SECTION
RECEIVED

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

COREY S WILSON

Signature and Printed Name

01/31/2017
Date