# Florida Department of Law Enforcement

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER

Agency: BAKER COUNTY SO Time of Inspection: 01:16

Date of Inspection: 12/24/2017

ZER 8000 Testino Description Serial Number: 80-001087

Check or Test	YES	370
Date and/or Time Adjusted	165	NO
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free # 0.05g/210L Test Test (g/210L) (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32316080A3 Exp: 01/05/2019
0.000	0.078	0.199	0.082
0.000	0.079	0.200	0.082
0.000	0.079	0.200	0.082

Number of Simulators Used: 5

Remarks:

LEC 27 2017

RECORDS SECTION

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

12/24/2017 Date

DEC 0 4 2017

RECEIVED

FDLE

#### AGENCY INSPECTION REPORT - INTOXILYZER 80 Tophol Testing Program

Agency: BAKER COUNTY SO Time of Inspection: 08:43

Date of Inspection: 11/26/2017

Serial Number: 80-001287

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
		No
Diagnostic Check (Pre-Inspection): OK		
	Yes	
Alcohol Free Subject Test: 0.000	-	
	Yes	
Mouth Alcohol Test: Slope Not Met		
	Yes	
Interferent Detect Test: Interferent Detect		
	Yes	
Diagnostic Check (Post-Inspection): OK		
	Yes	

Alcohol Free #0.05g/210L Test Test (g/210L) (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32316080A3 Exp: 01/05/2019
0.000	0.079	0.198	0.082
0.000 0.049	0.079	0.199	0.083
0.000 0.049	0.080	0.199	0.082

Number of Simulators Used: 5

Remarks:

1101 27 2011

COREY S WILSON

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

11/26/2017

Date

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO

Time of Inspection: 11:38

Date of Inspection: 10/29/2017

Serial Number: 80-001287

Software: 8100.27

Check or Test		
Date and/or Time Adjusted	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	
	Yes	

Alcohol Free	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32316080A3
0.000 0.049	0.079	0.197	Exp: 01/05/2019
0.000	0.079		0.082
0.000		0.198	0.082
0.000 [0.049	0.080	0.199	0.082

Number of Simulators Used: 5

Remarks:

Time-Date changed

BAKER COUNTY SHERIFF'S OFFICE

NOV 0 1 2017

RECORDS SECTION

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspects on in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

10/29/2017 Date

RECEIVED

OCT 05 2017

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: BAKER COUNTY SO Time of Inspection: 19:15

Date of Inspection: 09/30/2017

Serial Number: 80-001287 Software: 8100.27

Check or Test	VEC.	370
Date and/or Time Adjusted	YES	NO
and, or time Adjusted		
Diamenti di i		No
Diagnostic Check (Pre-Inspection): OK		
	Yes	
Alcohol Free Subject Test: 0.000		
	Yes	
Mouth Alcohol Test: Slope Not Met		_
	Yes	
Interferent Detect Test: Interferent Detect	100	
	Yes	
Diagnostic Check (Post-Inspection): OK	105	
	Yes	
	165	

Alcohol Free 0.05g/210L Test  Test (g/210L)  (g/210L)  Lot#:201603D  Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32316080A3 Exp: 01/05/2019
0.000	0.078	0.197	0.082
0.000 0.048	0.079	0.198	0.081
0.000	0.079	0.198	0.081

Number of Simulators Used: 5

Remarks:

BAKER COUNTY SHERIFF'S OFFICE

OCT 2 2017

RECORDS SECTION RECEIVED

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

09/30/2017 Date

### Florida Department of Law Enforcement Alcohol Testing Program

SEP 2 9 2017

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Testing Program

Agency: BAKER COUNTY SO Time of Inspection: 10:27

Date of Inspection: 08/31/2017

Serial Number: 80-001287

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D EExp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32316080A3 Exp: 01/05/2019
0.000	0.048	0.078	0.198	0.082
0.000	0.049	0.079	0.199	0.081
0.000	0.049	0.079	0.199	0.082

			200	
Mumber	of	Simulators	Tised. 5	

Remarks:

BAKER COUNTY SHERIEF'S OFFICE

AUG 3 1 2017

RECEIVED

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

COREY S WILS

Signature and Printed Name

08/31/2017 Date

### Florida Department of Law Enforcement Alcohol Testing Program

AUG 0 4 2017

FDLE Alcohol Testing Program

#### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO Time of Inspection: 20:41

Date of Inspection: 07/31/2017

Serial Number: 80-001287

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		140
		No
Diagnostic Check (Pre-Inspection): OK		
Algebel Free Orbins To accompany	Yes	
Alcohol Free Subject Test: 0.000		
Mouth Algohol Togt, Glans Web W.	Yes	
Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect	Yes	
Interferent Detect Test: Interferent Detect		
Diagnostic Check (Post-Inspection): OK	Yes	
s OK	1000	
The State of the Control of the State of the	Yes	

Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32316080A3 Exp: 01/05/2019
0.000	0.049	0.078	0.197	0.083
0.000	0.049	0.079	0.198	0.083
0.000	0.049	0.079	0.198	0.083

Number of Simulators Used: 5

Remarks:

BAKER COUNTY SHERIFF'S OFFICE

AUG 0 1 2017

RECEIVED

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

07/31/2017

Date

### Florida Department of Law Enforcement Alcohol Testing Program

JUL 0 5 2017

**FDLE** 

#### AGENCY INSPECTION REPORT - INTOXILYZER 800 Acohol Testing Program

Agency: BAKER COUNTY SO Time of Inspection: 18:58

Date of Inspection: 06/29/2017

Serial Number: 80-001287

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
		No
Diagnostic Check (Pre-Inspection): OK		
	Yes	
Alcohol Free Subject Test: 0.000		
	Yes	
Mouth Alcohol Test: Slope Not Met		
	Yes	
Interferent Detect Test: Interferent Detect		
	Yes	
Diagnostic Check (Post-Inspection): OK		
	Yes	

Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32316080A3 Exp: 01/05/2019
0.000	0.048	0.078	0.196	0.082
0.000	0.048	0.079	0.196	0.082
0.000	0.049	0.079	0.197	0.082

Number of Simulators Used: 5

Remarks:

BAKER COUNTY SHERIFF'S OFFICE

JUN 30 2017

RECORDS SECTION
RECEIVED

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

06/29/2017 Date

JUN 0 2 2017

FDLE

### AGENCY INSPECTION REPORT - INTOXILYZER 8000 Testing Programs

Agency: BAKER COUNTY SO Time of Inspection: 06:02

Date of Inspection: 05/28/2017

Serial Number: 80-001287

Sof	tware:	8100	. 27

Check or Test	YES	NO
Date and/or Time Adjusted		110
Diagnostic Check (Pre-Inspection): OK		No
	Yes	
Alcohol Free Subject Test: 0.000	165	1
Mouth Alcohol Test: Slope Not Met	Yes	
	Yes	
Interferent Detect Test: Interferent Detect	163	1
Diagnostic Check (Post-Inspection): OK	Yes	
- The state of the period of t	Yes	

Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32316080A3 Exp: 01/05/2019
0.000	0.048	0.077	0.196	0.082
0.000	0.048	0.078	0.198	0.082
0.000	0.049	0.079	0.198	0.082

Number of Simulators Used: 5

Remarks:

BAKER COUNTY SHERIFF'S OFFICE

MAY 3 0 2017

RECORDS SECTION
RECEIVED

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

COREY S WILSON

Signature and Printed Name

05/28/2017

Date



### Florida Department of Law Enforcement Alcohol Testing Program

MAY 0 9 2017

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: BAKER COUNTY SO Time of Inspection: 19:26

Date of Inspection: 04/20/2017

Serial Number: 80-001287

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	125	NO
Diagnostic Check (Pre-Inspection): OK		No
product theck (Fie-Inspection): OK		
Alcohol Free Subject Test: 0.000	Yes	
Mouth Algebel West Gl	Yes	
Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect	Yes	
	Yes	
Diagnostic Check (Post-Inspection): OK		
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32316080A3 Exp: 01/05/2019
0.000	0.048	0.078	0.195	0.082
0.000	0.048	0.079	0.196	0.082
0.000	0.049	0.079	0.197	0.081

Number of Simulators Used: 5

Remarks:

SHEAT COUNTY SHEAT PETCE APR 2 1 2017

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

COREY S V Signature and Printed Name

04/20/2017 Date

### Florida Department of Law Enforcement Alcohol Testing Program

APR 1 1 2017

FDLE

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO

Time of Inspection: 20:08

Date of Inspection: 03/31/2017

Serial Number: 80-001287

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		140
Diagnostic Check (Pre-Inspection): OK		No
	Yes	
Alcohol Free Subject Test: 0.000	100	
Mouth Alcohol Test: Slope Not Met	Yes	
	Yes	
Interferent Detect Test: Interferent Detect		
Diagnostic Check (Post-Inspection): OK	Yes	-
*	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:32316080A3 Exp: 01/05/2019
0.000	0.048	0.077	0.196	0.082
0.000	0.049	0.078	0.197	0.082
0.000	0.048	0.079	0.197	0.081

Number of Simulators Used: 5

Remarks:

EARTH COUNTY SHEET FOR TOTAL AAPR 7 702017 ECONDS SECTION

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

alle.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

COREY S WILSON

Signature and Printed Name

03/31/2017 Date

#### Florida Department of Law Enforcement FEB 07 2017 Alcohol Testing Program FDLE Alcohol Testing Program

#### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO

Time of Inspection: 01:10

Date of Inspection: 01/31/2017

Serial Number: 80-001287

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
		No
Diagnostic Check (Pre-Inspection): OK		
	Yes	
Alcohol Free Subject Test: 0.000		
	Yes	
Mouth Alcohol Test: Slope Not Met	- 11	
	Yes	
Interferent Detect Test: Interferent Detect		
	Yes	
Diagnostic Check (Post-Inspection): OK		
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A1 Exp: 10/05/2017
0.000	0.049	0.079	0.199	0.081
0.000	0.050	0.079	0.199	0.082
0.000	0.050	0.080	0.200	0.081

Number of Simulators Used: 5

Remarks:

RECORDS SECTION

Th	e above	instr	umen	t co	mpl	ies (	х )	does	not c	comply	• (	) with	Chapter	11D-8,	FAC.			
I	certify	that	I b	old	a	valid	Florid	a Dep	artme	ent of	Law	Enforcem	ent Agen	cy Insp	ector	Permit	and	that
De	erformed	this	Dasp	ecti	on	in acc	cordance	e with	the	provi	sions	of Chapt	er 11D-8	, FAC.				

Signature and Printed Name

01/31/2017 Date