





# Florida Department of Law Enforcement Alcohol Testing Program

JUN 20 2017

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAY COUNTY SO  
Time of Inspection: 12:25

Date of Inspection: 06/19/2017

Serial Number: 80-001309  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_

Remarks:  
TEST OF FORMS LOAD

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

PATRICK J MURPHY

Signature and Printed Name

06/19/2017  
Date



# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAY COUNTY SO  
Time of Inspection:

Date of Inspection: 04/13/2017

Serial Number: 80-001309  
Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000		
Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect		
Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: \_\_\_\_\_



Remarks:

UNABLE TO COMPLETE APRIL 2017 AGENCY INSPECTION DUE TO INSTRUMENT POWER ISSUES. THE INSTRUMENT WILL NOT POWER UP INTO READY MODE.

INSTRUMENT WILL BE TAKEN OUT OF SERVICE FOR EVIDENTIARY USE ON 04/13/2017 TO BE SENT IN FOR REPAIRS.

The above instrument complies (\_\_\_\_) does not comply (\_\_\_\_) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


  
 Signature and Printed Name