

80-001345

Florida Department of Law Enforcement

Alcohol Testing Program

RECEIVED

DEC 22 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: EUSTIS P.D.

Time of Inspection: 08:38

Date of Inspection: 12/22/2017

Serial Number: 80-001345

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 203498 Exp: 12/10/2018
0.000	0.049	0.078	0.194	0.078
0.080	0.049	0.078	0.194	0.078
0.000	0.049	0.078	0.195	0.078

Number of Simulators Used: 2

Remarks: MERRY CHRISTMAS!
HAPPY HOLIDAYS!
HAPPY NEW YEAR!
SEASON'S GREETINGS!

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ROBERT T. SIMKEN

Signature and Printed Name

12/22/2017

Date

80-001345 *ENG*

RECEIVED

NOV 21 2017

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: EUSTIS P.D.
Time of Inspection: 08:58

Date of Inspection: 11/20/2017

Serial Number: 80-001345
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 203498 Exp: 12/10/2018
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.079	0.198	0.078

Number of Simulators Used: 5

Remarks:

Int. Dev: SIM LOOSE.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

ROBERT T SIMKEN

11/20/2017
Date

80-001345 DB

Florida Department of Law Enforcement

Alcohol Testing Program

RECEIVED

OCT 27 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: EUSTIS P.D.

Time of Inspection: 11:51

Date of Inspection: 10/27/2017

Serial Number: 80-001345

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 2016030 Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 203498 Exp: 12/10/2018
0.000	0.048	0.079	0.198	0.079
0.000	0.048	0.078	0.199	0.078
0.000	0.048	0.078	0.198	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DET Robert T. Sinken

ROBERT T. SINKEN

Signature and Printed Name

10/27/2017

Date

80-001345 ^{EMB}

Florida Department of Law Enforcement

Alcohol Testing Program

RECEIVED

SEP 05 2017

AGENCY INSPECTION REPORT - INTOXILYZER 8000

FDLE

Alcohol Testing Program

Agency: EUSTIS P.D.

Time of Inspection: 10:08

Date of Inspection: 09/05/2017

Serial Number: 80-001345

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 203498 Exp: 12/10/2018
0.000	0.048	0.078	0.197	0.079
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.079	0.198	0.078

Number of Simulators Used: 5

Remarks:



The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

ROBERT T SIMKEN

09/05/2017
Date

Florida Department of Law Enforcement

Alcohol Testing Program

AGENCY INSPECTION REPORT -- INTOXILYZER 8000

Agency: EUSTIS P.D.

Time of Inspection: 09:54

Date of Inspection: 08/29/2017

Serial Number: 80-001345

Software: 8100.27

RECEIVED

AUG 29 2017

FDLE
Alcohol Testing Program

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 2016030 Exp: 03/06/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 203498 Exp: 12/10/2018
0.000	0.048	0.078	0.198	0.078
0.000	0.048	0.078	0.199	0.078
0.000	0.048	0.078	0.199	0.078

Number of Simulators Used: 3

Remarks:



The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DET Robert T. Simken

ROBERT T. SIMKEN

Signature and Printed Name

08/29/2017
Date

80-001345 RMB

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
JUL 24 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: EUSTIS P.D.
Time of Inspection: 09:39

Date of Inspection: 07/24/2017

Serial Number: 80-001345
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 203498 Exp: 12/10/2018
0.000	0.048	0.077	0.196	0.078
0.000	0.048	0.078	0.196	0.078
0.000	0.049	0.079	0.196	0.078

Number of Simulators Used: 5

Remarks:

CLOCK RESET
CLOCK RPTD BY E42
AS BEING INCORRECT
POWER OUTAGES IN AREA
DL

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ROBERT T. SIMKEN

Signature and Printed Name

07/24/2017
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
JUN 26 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: EUSTIS P.D.

Time of Inspection: 10:00

Date of Inspection: 06/26/2017

Serial Number: 80-001345

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 203498 Exp: 12/10/2018
0.000	0.049	0.078	0.197	0.079
0.000	0.049	0.079	0.196	0.078
0.000	0.049	0.078	0.197	0.078

Number of Simulators Used: 5

Remarks:



The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Det Robert T. Shiken

ROBERT T. SHIKEN

Signature and Printed Name

06/26/2017
Date

80-001345 *QMS*

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
MAY 08 2017
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: EUSTIS P.D.

Time of Inspection: 09:46

Date of Inspection: 05/08/2017

Serial Number: 80-001345

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 203498 Exp: 12/10/2013
0.000	0.048	0.079	0.196	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.079	0.197	0.079

Number of Simulators Used: 5

Remarks:

Int Det: LID LOOSE *ON 1ST ATTEMPT RS*The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DET Robert T. Simken

ROBERT T. SIMKEN

Signature and Printed Name

05/08/2017
Date

80-001345 *FWB*

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

APR 26 2017

FDLE
Alcohol Testing Program**AGENCY INSPECTION REPORT - INTOXILYZER 8000**

Agency: EUSTIS P.D.

Time of Inspection: 09:31

Date of Inspection: 04/26/2017

Serial Number: 80-001345

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 203498 Exp: 12/10/2018
0.000	0.049	0.079	0.195	0.079
0.000	0.050	0.079	0.196	0.079
0.000	0.049	0.080	0.196	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

ROBERT T. SIMKEN

Signature and Printed Name

04/26/2017

Date

80-001345 RAS

Florida Department of Law Enforcement

Alcohol Testing Program

RECEIVED

APR 03 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: EUSTIS P.D.

Time of Inspection: 18:54

Date of Inspection: 03/27/2017

Serial Number: 80-001345

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 203498 Exp: 12/10/2018
0.000	0.048	0.076	0.196	0.078
0.000	0.049	0.078	0.197	0.078
0.000	0.049	0.077	0.197	0.078


Number of Simulators Used: 5

Remarks:

NONE

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sr. K9 CPB  F36

Signature and Printed Name

MATTHEW J. CRIEVES

03/27/2017
Date

80-001345

Florida Department of Law Enforcement

Alcohol Testing Program

RECEIVED

MAR 28 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: EUSTIS P.D.

Time of Inspection: 18:54

Date of Inspection: 03/27/2017

Serial Number: 80-001345

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 203498 Exp: 12/10/2018
0.000	0.048	0.076	0.196	0.078
0.000	0.049	0.078	0.197	0.078
0.000	0.049	0.077	0.197	0.078

Number of Simulators Used: 5

Remarks:

NONE

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Sr. K9 CR *[Signature]* F36

MATTHEW J GRIEVES

Signature and Printed Name

03/27/2017
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

APR 03 2017

**FDLE
Alcohol Testing Program****AGENCY INSPECTION REPORT - INTOXILYZER 8000**

Agency: BUSTIS P.D.

Time of Inspection: 18:15

Date of Inspection: 03/27/2017

Serial Number: 80-001345

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Samples Used: 5

Remarks:

A.P.D.M.

Introduced mouth alcohol on alcohol free sample.

Inspection restarted.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

S. A. [Signature]

Signature and Printed Name

MATTHEW J. GRIEVED

03/27/2017
Date

80-001345

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

MAR 28 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BUSTIS P.D.

Time of Inspection: 18:16

Date of Inspection: 03/27/2017

Serial Number: 80-001345

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators used: 0

Remarks:

APFAA:

Introduced mouth alcohol on alcohol free sample.

Inspection restarted.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

S. A. 

MATTHEW J. GRIEVES

Signature and Printed Name

03/27/2017
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
FEB 24 2017
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: EUSTIS P.D.

Time of Inspection: 09:52

Date of Inspection: 02/24/2017

Serial Number: 80-001345

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 203498 Exp: 12/10/2018
0.000	0.049	0.078	0.196	0.079
0.000	0.049	0.078	0.197	0.079
0.000	0.050	0.079	0.198	0.079

Number of Simulators Used: 5

Remarks: **IN COMPLIANCE** 😊

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Det Robert T Simken
Signature and Printed Name

02/24/2017
Date

80-001345

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JAN 24 2017

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: EUSTIS P.D.

Time of Inspection: 09:21

Date of Inspection: 01/24/2017

Serial Number: 80-001345

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:203498 Exp: 12/10/2018
0.000	0.050	0.078	0.197	0.079
0.000	0.051	0.078	0.197	0.078
0.000	0.051	0.079	0.199	0.078

Number of Simulators Used: 5

Remarks: IN COMPLIANCE

RETURNED FROM FDLE - BACK IN SERVICE.

RS

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DET Robert T. SIMKEN

ROBERT T. SIMKEN

Signature and Printed Name

01/24/2017

Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JAN 13 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: EUSTIS P.D.

Time of Inspection: 09:59

Date of Inspection: 01/13/2017

Serial Number: 80-001345

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:203498 Exp: 12/10/2018
0.000	0.048	0.078	0.199	0.080
0.000	0.048	0.078	0.200	0.080
0.000	0.048	0.079	0.199	0.080

Number of Simulators Used: 5

Remarks: SENT TO FDLE FOR ANNUAL INSPECTION

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DET R. R. S. - ERY

ROBERT T SIMKEN

Signature and Printed Name

01/13/2017

Date

RECEIVED

Florida Department of Law Enforcement

Alcohol Testing Program

JAN 24 2017

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: EUSTIS P.D.

Time of Inspection: 09:21

Date of Inspection: 01/24/2017

Serial Number: 80-001345

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:203498 Exp: 12/10/2018
0.000	0.050	0.078	0.197	0.079
0.000	0.051	0.078	0.197	0.078
0.000	0.051	0.079	0.199	0.078

Number of Simulators Used: 5


Remarks: IN COMPLIANCE

RETURNED FROM FDLE - BACK IN SERVICE.



The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ROBERT T. SIMKEN

Signature and Printed Name

01/24/2017
Date