

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

OCT 02 2017

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: ORANGE COUNTY S.O.
Time of Inspection: 12:31

Date of Inspection: 09/21/2017

Serial Number: 80-001416
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#: 201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: 201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 26316080A1 Exp: 11/05/2018
0.000	0.048	0.195 / 0.078	0.194	0.080
0.000	0.048	0.197 / 0.078	0.196	0.080
0.000	0.049	0.198 / 0.079	0.196	0.080

Number of Simulators Used: 5


Remarks:

08: Control Outside Tolerance ATTCHD .20 SIM INSTEAD OF .08. .08 ARS LOT NO. 201611B EXP 11/15/2018

Repeated .08 Test with correct simulator attached. *dm*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



KELLY M MELVILLE

Signature and Printed Name

09/21/2017
Date