

RECEIVED

Florida Department of Law Enforcement
Alcohol Testing Program

AUG 14 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COLLIER COUNTY SO
Time of Inspection: 12:10

Date of Inspection: 08/03/2017

Serial Number: 80-001651
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:34015080A2 Exp: 01/05/2018
/ 0.000	0.048	0.077	0.194	0.078
/ 0.000	0.048	0.079	0.196	0.079
/ 0.000	0.049	0.078	0.196	0.079

Number of Simulators Used: 5

Remarks:

A F / M A:TOO MUCH MOUTH ALCOHOL. 00: Ambient Fail.

Residual mouth alcohol on hands, hand was resting to close breath tube opening. KLB#488

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Kelly Bradley #488

KELLY L BRADLEY
Signature and Printed Name

08/03/2017
Date