

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

RECEIVED  
SEP 06 2017  
FDLE  
Alcohol Testing Program

Agency: St Petersburg Police Department  
Serial Number: 80-001653  
Time of Inspection: 08/30/2017  
Software:

Date of Inspection: 08/30/22017

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000		
Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect		
Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: \_\_\_\_\_

Remarks:

Unable to perform Inspection Test due to instrument not reading control test.

The above instrument complies ( ) does not comply ( XXX ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Michael Weiskopf*

Michael Weiskopf

Signature and Printed Name

Date

ST PETERSBURG PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001653  
08/30/2017  
Software: 8100.27

RECEIVED  
SEP 06 2017  
FDLE  
Alcohol Testing Program

-----  
DRY CONTROL TEST

Test	g/210L	Time
Air Blank	0.000	08:04
Control Test	0.000	08:05
Air Blank	0.000	08:05

  
Operator's Signature

ST PETERSBURG PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001653  
08/30/2017  
Software: 8100.27

RECEIVED  
SEP 06 2017  
FDLE  
Alcohol Testing Program

-----  
DRY CONTROL TEST

Test	g/210L	Time
Air Blank	0.000	08:03
Control Test	0.000	08:03
Air Blank	0.000	08:03

  
Operator's Signature

ST PETERSBURG PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001653  
08/30/2017  
Software: 8100.27

RECEIVED  
SEP 06 2017  
FDLE  
Alcohol Testing Program

-----  
ITP CONTROL TEST

Test	g/210L	Time
Air Blank	0.000	07:01
ITP Check	0.080	07:01
Air Blank	0.000	07:02

  
Operator's Signature

ST PETERSBURG PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001653  
08/30/2017  
Software: 8100.27

**RECEIVED**  
SEP 08 2017  
FDLE  
Alcohol Testing Program

-----  
DRY CONTROL TEST

Test	g/210L	Time
Air Blank	0.000	07:02
Control Test	0.000	07:03
Air Blank	0.000	07:03

  
Operator's Signature

This is the test when the problem occurred

FLORIDA DEPARTMENT OF LAW ENFORCEMENT  
ALCOHOL TESTING PROGRAM  
BREATH ALCOHOL TEST AFFIDAVIT

RECEIVED  
SEP 06 2017  
FDLE  
Alcohol Testing Program

Instrument Type: Intoxilyzer 8000  
Instrument Registered To: ST PETERSBURG PD  
Instrument Serial Number: 80-001653 Software: 8100.27  
Date of Test: 08/30/2017

Date of Last Agency Inspection: 08/05/2017  
Observation Period Began: 01:33  
Subject's Name: CHRISTOPHER J BAIETTO

DOB: 07/29/1989 Sex: M

The subject was observed for at least twenty-minutes prior to the administration of the breath test to ensure that the subject did not take anything orally and did not regurgitate.

Results:	Test	g/210L	Time
	Diagnostics Check	OK	01:58
	Air Blank	0.000	01:58
	Control Test	0.000*	01:59
	Air Blank	0.000	01:59

\*Control Outside Tolerance

Cylinder Lot: 28515080A2  
Exp: 11/05/2017

State of Florida, County of \_\_\_\_\_,

Personally appeared before me the undersigned authority, who (\_\_\_) is personally known to me or (\_\_\_) produced \_\_\_\_\_ as identification, and who after being placed under oath, states:

I ANTHONY ALLE, hold a valid Breath Test Operator permit issued by the Florida Department of Law Enforcement, I administered the above breath test to the subject named above in accordance with Chapter 11D-8, Florida Administrative Code, and this form is a true and accurate report of that breath test.

Breath Test Operator: \_\_\_\_\_ Date: \_\_\_\_\_  
Signature

Sworn to (or affirmed) before me this \_\_\_\_\_ day of \_\_\_\_\_, \_\_\_\_\_

Signature of Notary Public-State of Florida \_\_\_\_\_ Printed Name of Notary Public-State of Florida \_\_\_\_\_

Note: Pursuant to section 117.10, Florida Statutes, law enforcement officers, correctional officers, traffic accident investigation officers and traffic infraction enforcement officers are notaries public when engaged in the performance of official duties. In accordance with section 316.1934(5), F.S., this completed form is admissible without further authentication and is presumptive proof of the results herein. To be used in accordance with Section 316.1934(5), F.S., and in administrative proceedings pursuant to 322.2615, F.S.

FLORIDA DEPARTMENT OF LAW ENFORCEMENT  
ALCOHOL TESTING PROGRAM  
BREATH ALCOHOL TEST AFFIDAVIT

RECEIVED  
SEP 06 2017  
FDLE  
Alcohol Testing Program

Instrument Type: Intoxilyzer 8000  
Instrument Registered To: ST PETERSBURG PD  
Instrument Serial Number: 80-001653 Software: 8100.27  
Date of Test: 08/30/2017

Date of Last Agency Inspection: 08/05/2017

Observation Period Began: 01:33

Subject's Name: CHRISTOPHER J BAIETTO

DOB: 07/29/1989 Sex: M

The subject was observed for at least twenty-minutes prior to the administration of the breath test to ensure that the subject did not take anything orally and did not regurgitate.

Results:	Test	g/210L	Time
	Diagnostics Check	OK	01:53
	Air Blank	0.000	01:54
	Control Test	0.000*	01:54
	Air Blank	0.000	01:55

\*Control Outside Tolerance

Cylinder Lot: 28515080A2  
Exp: 11/05/2017

State of Florida, County of \_\_\_\_\_,

Personally appeared before me the undersigned authority, who ( ) is personally known to me or ( ) produced \_\_\_\_\_ as identification, and who after being placed under oath, states:

I ANTHONY ALLE, hold a valid Breath Test Operator permit issued by the Florida Department of Law Enforcement, I administered the above breath test to the subject named above in accordance with Chapter 11D-8, Florida Administrative Code, and this form is a true and accurate report of that breath test.

Breath Test Operator: \_\_\_\_\_ Signature \_\_\_\_\_ Date: \_\_\_\_\_

Sworn to (or affirmed) before me this \_\_\_\_\_ day of \_\_\_\_\_, \_\_\_\_\_

Signature of Notary Public-State of Florida \_\_\_\_\_ Printed Name of Notary Public-State of Florida \_\_\_\_\_

Note: Pursuant to section 117.10, Florida Statutes, law enforcement officers, correctional officers, traffic accident investigation officers and traffic infraction enforcement officers are notaries public when engaged in the performance of official duties. In accordance with section 316.1934(5), F.S., this completed form is admissible without further authentication and is presumptive proof of the results herein. To be used in accordance with Section 316.1934(5), F.S., and in administrative proceedings pursuant to 322.2615, F.S.