

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

APR 19 2017

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 10:19

Date of Inspection: 04/19/2017

Serial Number: 80-003347
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A2 Exp: 10/05/2017
/ 0.000	0.049	0.077	0.197	0.078
/ 0.000	0.050	0.079	0.199	0.078
/ 0.000	0.051	0.079	0.199	0.078

Number of Simulators Used: 5

Remarks:
A F / M A: RFI Detect. 00: Ambient Fail.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

ADAM D CINELLI

Signature and Printed Name

04/19/2017
Date

80-003347 RMB

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003347
04/19/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:27
Control Test	0.077	09:28
Air Blank	0.000	09:28
Control Test	0.078	09:29
Air Blank	0.000	09:30
Control Test Stats		
Average	0.0775	
Std Dev	0.0007	
Rel Std Dev(%)	0.9124	



Operator's Signature

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003347
04/19/2017
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:32
Control Test	0.078	09:33
Air Blank	0.000	09:33
Control Test	0.079	09:34
Air Blank	0.000	09:34
Control Test Stats		
Average	0.0785	
Std Dev	0.0007	
Rel Std Dev(%)	0.9008	



Operator's Signature

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

APR 19 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO

Time of Inspection: 08:51

Date of Inspection: 04/19/2017

Serial Number: 80-003347

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____
0.000	0.049	0.074 / 0.071		
0.000	0.049	0.074 / 0.071		
0.000	0.049	0.075 / 0.072		

Number of Simulators Used: 5


Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

** CONTACTED FDLE VIA PHONE.
CONDUCTED STABILITY CHECKS AFTER CHECKING SIMULATOR CONNECTIONS AND
FOUND RESULTS WITHIN RANGE. RE-DO INSPECTION;*

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ADAM D CINELLI

Signature and Printed Name

04/19/2017
Date