80-004587 2018

RECEIVED
DEC 02 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY S.O.
Time of Inspection: 21:04 Date of Inspection: 12/01/2017 Serial Number: 80-004587
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:163437 Exp: 10/30/2018
0.000	0.046	0.076	0.196	0.078
0.000	0.047	0.077	0.198	0.078
0.000	0.047	0.078	0.199	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D. Lyons
DANIEL E LYONS
Signature and Printed Name

12/01/2017
Date

FDLE/ATP Form 40 – March 2004

80-004587 2018

RECEIVED
NOV 07 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY S.O.
Time of Inspection: 18:52 Date of Inspection: 11/07/2017 Serial Number: 80-004587
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:163437 Exp: 10/30/2018
0.000	0.047	0.077	0.193	0.079
0.000	0.047	0.077	0.195	0.078
0.000	0.048	0.078	0.195	0.078

Number of Simulators Used: 5

Remarks:
Time-Date changed.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D Lyons
DANIEL E LYONS
Signature and Printed Name

11/07/2017
Date

FDLE/ATP Form 40 – March 2004

80-004587 DAB

RECEIVED
OCT 19 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY S.O.
Time of Inspection: 21:42 Date of Inspection: 10/18/2017 Serial Number: 80-004587
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:163437 Exp: 10/30/2018
0.000	0.048	0.077	0.194	0.080
0.000	0.048	0.078	0.196	0.079
0.000	0.047	0.078	0.196	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D. Lyons DANIEL E LYONS
Signature and Printed Name

10/18/2017
Date

FDLE/ATP Form 40 – March 2004

80-004587 KDB

RECEIVED
SEP 28 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY S.O.
Time of Inspection: 17:44 Date of Inspection: 09/28/2017
Serial Number: 80-004587
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:163437 Exp: 10/30/2018
0.000	0.046	0.077	0.194	0.079
0.000	0.046	0.077	0.195	0.079
0.000	0.047	0.078	0.195	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D Lyons

DANIEL E LYONS
Signature and Printed Name

09/28/2017
Date

FDLE/ATP Form 40 – March 2004

80-004587 OSCS

RECEIVED
AUG 29 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY S.O.
Time of Inspection: 00:01 Date of Inspection: 08/29/2017
Serial Number: 80-004587
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.046	0.074 / 0.070		
0.000	0.047	0.074 / 0.073		
0.000	0.047	0.074 / 0.074		

Number of Simulators Used: 5

Remarks:
Int Det: . 08: Control Outside Tolerance olerance. .08 SOLUTION OUT OF TOLERANCE.

The above instrument complies ( ) does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D Lyons
DANIEL E LYONS
Signature and Printed Name

08/29/2017
Date

FDLE/ATP Form 40 – March 2004

80-004587

RECEIVED
AUG 29 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY S.O.
Time of Inspection: 21:18 Date of Inspection: 08/28/2017
Serial Number: 80-004587
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.046	0.072 / 0.072		
0.000	0.046	0.074 / 0.074		
0.000	0.046	0.073 / 0.075		

Number of Simulators Used: 5

Remarks:
08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

The above instrument complies ( ) does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D. Lyons
DANIEL E LYONS
Signature and Printed Name

08/28/2017
Date

FDLE/ATP Form 40 – March 2004

80-004587 RWS

RECEIVED
AUG 02 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY S.O.
Time of Inspection: 12:15 Date of Inspection: 08/01/2017
Serial Number: 80-004587
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:20163D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:163437 Exp: 10/30/2018
0.000	0.048	0.076	0.193	0.079
0.000	0.048	0.076	0.193	0.078
0.000	0.048	0.077	0.194	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D. Lyons

DANIEL E LYONS
Signature and Printed Name

08/01/2017
Date

FDLE/ATP Form 40 – March 2004

80-004587 DMS

RECEIVED
JUL 11 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY S.O.
Time of Inspection: 15:42 Date of Inspection: 07/11/2017 Serial Number: 80-004587
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:163437 Exp: 10/30/2018
0.000	0.048	0.077	0.194	0.079
0.000	0.048	0.077	0.196	0.079
0.000	0.047	0.077	0.195	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D Lyons

DANIEL E LYONS
Signature and Printed Name

07/11/2017
Date

FDLE/ATP Form 40 – March 2004

80-004587 RWB

RECEIVED
JUN 28 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY S.O.
Time of Inspection: 18:40 Date of Inspection: 06/27/2017
Serial Number: 80-004587
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:163437 Exp: 10/30/2018
0.000	0.046	0.075	0.194	0.079
0.000	0.047	0.077	0.195	0.078
0.000	0.047	0.077	0.195	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DANIEL E LYONS
Signature and Printed Name

06/27/2017
Date

FDLE/ATP Form 40 – March 2004

80-004587

RECEIVED
MAR 02 2017
FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY S.O.
Time of Inspection: 06:03
Date of Inspection: 02/15/2017
Serial Number: 80-004587
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 63437 Exp: 10/30/2018
0.000	0.047	0.075	0.190	0.079
0.000	0.048	0.077	0.193	0.078
0.000	0.048	0.076	0.194	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

ANGEL M DIAZ
Signature and Printed Name

02/15/2017
Date

FDLE/ATP Form 40 - March 2004