

# Florida Department of Law Enforcement Alcohol Testing Program

101108  
SP

**RECEIVED**

APR 20 2017

**FDLE**

## AGENCY INSPECTION REPORT – INTOXILYZER 8000 Alcohol Testing Program

Agency: MARTIN COUNTY SO  
Time of Inspection: 12:10

Date of Inspection: 03/28/2017

Serial Number: 80-006168  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:20016080A2 Exp: 09/05/2018
0.000	0.049	0.079	0.198	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.050	0.079	0.199	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



BRADFORD B BAKER

Signature and Printed Name

03/28/2017  
Date