

6170 SR

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

JUN 01 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MARTIN COUNTY SO
Time of Inspection: 13:45

Date of Inspection: 05/26/2017

Serial Number: 80-006170
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:31715080A3 Exp: 12/05/2017
0.000	0.049	0.059 / 0.077	0.198	0.076
0.000	0.049	0.000 / 0.077	0.198	0.079
0.000	0.049	0.000 / 0.077	0.198	0.079

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance. SIM BAD, CHANGED AND RETESTED

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D/S C. [Signature] 5/11/1602 CHRISTOPHER L THOMAS
Signature and Printed Name

05/26/2017
Date