80-001e170

Florida Department of Law Enforcement **Alcohol Testing Program**

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MARTIN COUNTY SO Time of Inspection: 13:45

Date of Inspection: 05/26/2017

Serial Number: 80-006170

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
		No
Diagnostic Check (Pre-Inspection): OK		
	Yes	
Alcohol Free Subject Test: 0.000		
	Yes	
Mouth Alcohol Test: Slope Not Met		
	Yes	
Interferent Detect Test: Interferent Detect		
	Yes	
Diagnostic Check (Post-Inspection): OK		1
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:31715080A3 Exp: 12/05/2017	
0.000	0.049	0.059 / 0.077	0.198	0.076	
0.000	0.049	0.000 / 0.077	0.198	0.079	
0.000	0.049	0.000 / 0.077	0.198	0.079	

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance. SIM BAD, CHANGED AND RETESTED S-26-17

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

1602 CHRISTOPHER L THOMAS Signature and Printed Name

05/26/2017



AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Martin County SO Instrument Serial Number: 80-006170

AGENCY INSPECTION DATA REVIEW					
Agency Inspector: Christopher Thomas		Date of Inspection: 5/26/1	7	Time of Inspection: 13:45	
		1		•	
		Untimely/Not Received [] Erroneo	us Information	
		Other			
☐ Agency Inspection Not Conducted or I	Records regarding A	gency inspection have not b	een uploade	d.	
☐ Lot Number ☐ Expiration Date for		Reference Solution □Dry Ga		100 - 100 -	
 ✓ FDLE/ATP Form 39 states in part, "If a Remarks section of FDLE/ATP Form 4 test(s); OR the ✓ Possible Cause and ✓ Alcohol Free Subject Test 	0 Agency Inspection I Corrective Action T	Report – Intoxilyzer 8000. Taken on the following test(s)	he □REASO was not rec	ON for repeating the following	
□ 0.05 g/210L Test ⊠	0.08 g/210L Test	☐ 0.20 g/210L Test	000000000000000000000000000000000000000	3 g/210L Dry Gas Standard Test	
□ FDLE/ATP Form 39 states in part, "If the instrument from service and notify the □ The Department Inspector was not Inspection complies with the red □ The Department Inspector was not Chapter 11D-8, FAC and the interest	Department Inspect ot notified. However juirements of Chapte ot notified. However instrument was correct ot notified. The repe	or." , the issue was satisfactorily r 11D-8, FAC. , the repeated Agency Inspe ttly removed from evidentiar ated Agency Inspection doe	corrected and ction does not give use.	nd the repeated Agency ot comply with the requirements	
☐ The Agency Inspection is noted as "Co	omplies" when it does	s not comply with the require	ements of Ch	napter 11D-8, FAC.	
☐ Other:		000 800 00			
OTHER ELECTRONIC DATA REVIEW					
Login Records Date:	Comments:				
☐ Cylinder Change Records Date:					
☐ Control Test Records Date:					
☐ Diagnostic Check Records Date:					
		* * * · · · · · · · · · · · · · · · · ·			
CORRECTIVE ACTION					
Record hand-written amendments on the report "AMENDED", and forward a cop			, initial and d	late the amendments, mark the	
☐ Provide a written explanation regarding	the referenced item	(s) to the Department Inspec	tor by	(Date).	
☐ Upload the Agency Inspection(s).					
Remove the instrument from evidentiary use until otherwise directed by the Department.					
□ No action required □ Other:					
Shayla Platt			6/8/2017		
Signature of Alcohol Testing Program St	att Member		Date		

Florida Department of Law Enforcement **Alcohol Testing Program**

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

Agency MARTIN COUNTY SO

Time 13:45:09

Date 05/26/2017

Agency Inspection

Inspector THOMAS, CHRISTOPHER, L

Type Software

8100.27

0.05 Lot # 201603D

0.08 Lot # 201611B

0.20 Lot# 201604C

0.08 Gas Lot # 31715080A3

Expiration 03/08/2018

Expiration 11/15/2018

Expiration 04/05/2018

Expiration

12/05/2017

Number of Simulators 5

Compliance Yes

Check or Test	Result.	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.059
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.000
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.000
		0.08 g/210L Test # (Repeat) 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
Remarks:			

08: Control Outside Tolerance.

Data Download Date/Time 05/26/2017 13:45

Printed: 08 June 2017

COBRA ©1997-2014 CMI, Inc.



Florida Department of Law Enforcement JUN 0 1 2017 Alcohol Testing Program

FDLE **Alcohol Testing Program**

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MARTIN COUNTY SO Time of Inspection: 13:45

Date of Inspection: 05/26/2017

Serial Number: 80-006170

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
		No
Diagnostic Check (Pre-Inspection): OK		
	Yes	
Alcohol Free Subject Test: 0.000		
	Yes	
Mouth Alcohol Test: Slope Not Met		
	Yes	
Interferent Detect Test: Interferent Detect		
	Yes	
Diagnostic Check (Post-Inspection): OK		
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:31715080A3 Exp: 12/05/2017
0.000	0.049	0.059 / 0.077	0.198	0.076
0.000	0.049	0.000 / 0.077	0.198	0.079
0.000	0.049	0.000 / 0.077	0.198	0.079

Jumber	of	Simulatore	Head.	5	

Remarks:

08: Control Outside Tolerance. 5=M 3AD, CHANGED LND RETESTED

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I

performed this inspection in accordance with the provisions of Chapter 11D-8, FAC. 1602

CHRISTOPHER L THOMAS

Signature and Printed Name