Florida Department of Law Enforcement **Alcohol Testing Program**

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO

Time of Inspection: 09:58

Date of Inspection: 11/27/2017

Serial Number: 80-006236 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
		No
Diagnostic Check (Pre-Inspection): OK	400	
	Yes	
Alcohol Free Subject Test: 0.000	İ	-
	Yes	
Mouth Alcohol Test: Slope Not Met		T
	Yes	
Interferent Detect Test: Interferent Detect		
	Yes	
Diagnostic Check (Post-Inspection): OK		1
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201611B Exp: 11/15/2018	0.20g/210L Test (g/210L) Lot#:201702 Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:256835 Exp: 01/11/2019
0.000	0.048	0.076	0.194	0.082
0.000	0.048	0.076	0.194	0.081
0.000	0.048	0.077	0.195	0.081

Number of Simulators Used: 5

Remarks:

TROOPER LEDFORD PRESENT

COTTECHE LOTH FOR 2.209/210 + +PS+ (201702B)

The	above	instrument	complies	X)	does	not	comply	() with	Chapter	11D-8,	FAC.
-----	-------	------------	----------	---	---	------	-----	--------	---	--------	---------	--------	------

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

GLENN CLINE

Signature and Printed Name

11/27/2017

RECEIVED

Florida Department of Law Enforcement AUG 2 3 2017 Alcohol Testing Program

FDLE

No

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Testing Program

Agency: MANATEE COUNTY SO Time of Inspection: 09:13

Date of Inspection: 08/22/2017

Serial Number: 80-006236 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted Diagnostic Check (Pre-Inspection): OK Diagnostic Subject Test: 0.000 Diagnostic Slope Not Met		
		No
Diagnostic Check (Pre-Inspection): OK		
		No
Alcohol Free Subject Test: 0.000		
		No
Mouth Alcohol Test: Slope Not Met		
		No
Interferent Detect Test: Interferent Detect		
		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

ишшрег	OI	Simulators	used:	
Remarks	3:			

BYPASSED AI TO OPERATE INSTRUMENT

Diagnostic Check (Post-Inspection): OK

NIA	con	PLIANCE MINIED
1-1	NOT	DETERMINED

100				
The above instrument complies () does not comply () with Chapter 11D-8, FAC.				
I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector				
performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.	Permit	and	that	Ι
The provide the provide of chapter 11D-8, FAC.				
SHAYLA D PLATT	7			
Signature and Printed Name				

08/22/2017