

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

JUN 08 2017

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TROOP L  
Time of Inspection: 08:19

Date of Inspection: 04/10/2017

Serial Number: 80-006762  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

*ADM*

Number of Simulators Used: \_\_\_\_\_

Remarks:

COMPLIANCE NOT DETERMINED AI NOT CONDUCTED DISSABLED MODE

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*David E Reyes-Rivera*

DAVID E REYES-RIVERA

Signature and Printed Name

04/10/2017  
Date

*JL*