

RECEIVED

Florida Department of Law Enforcement
Alcohol Testing Program

APR 10 2017

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP

Time of Inspection: 21:11

Date of Inspection: 04/03/2017

Serial Number: 80-006768

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:34416080A2 Exp: 02/05/2019
0.000	0.045	0.075	0.195	0.079
0.000	0.046	0.075	0.195	0.079
0.000	0.046	0.076	0.196	0.079

Number of Simulators Used: 5

Remarks:

A F / M A: TOO MUCH MA

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

SUSAN M BARGE

Signature and Printed Name

04/03/2017

Date