

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

JUN 08 2017

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TROOP L

Time of Inspection: 08:03

Date of Inspection: 04/10/2017

Serial Number: 80-006769

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Handwritten initials

Number of Simulators Used: _____

Remarks:

COMPLIANCE NOT DETERMINED AI NOT CONDUCTED DISSABLED MODE

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Handwritten signature of David E Reyes-Rivera

DAVID E REYES-RIVERA

Signature and Printed Name

04/10/2017
Date

Handwritten initials