



**Bell, Danielle**

---

**From:** Collito, Brad <bcollito@hernandosheriff.org>  
**Sent:** Thursday, April 13, 2017 2:43 AM  
**To:** Bell, Danielle; Cameron, John IV; Conrad Juergensmeyer; Hutchinson, Harold; Ivette Miranda; Bell, Danielle; Page, Mike; Paula Tyrell; Sonya Wolthuis; Stevens, Teresa  
**Subject:** April 2017 Agency Inspections  
**Attachments:** April 2017 Agency Inspections.pdf

Good Morning,

Attached are the April 2017 Agency Inspections for the Hernando County Sheriff's Office.

Instrument 814 and 5249 were not inspected due to both of them being at CMI for servicing.

Instrument 228 is an Evidential Loaner instrument that we have received from FDLE while 814 and 5249 are being serviced.

Instrument 812 and 228 were inspected with no issues to report. Both instruments passed inspection.

It should be noted that since the repair of the electronic sliding door leading to the Sally Port, there have been no additional RFI Detects during tests and inspections.

If I can be of additional assistance, please do not hesitate to contact me.

Respectfully,  
Sgt. B. Collito

**Sergeant Brad Collito**  
**Intake/ Release Unit**  
**Hernando County Detention Center**  
**16425 Spring Hill Dr**  
**Brooksville, FL 34604**  
**T(352)544-2334 F(352)544-2349**  
**bcollito@hernandosheriff.org**

Please note: Florida has a very broad Public Records Law. Most written communications to or from the Hernando County Sheriff's Office are public records available to anyone upon request. Your e-mail, including your e-mail address, may therefore, be subject to public disclosure.

**Bell, Danielle**

80-000228

AB

80-000812

80-000814

80-005249

**From:** Collito, Brad <bcollito@hernandosheriff.org>  
**Sent:** Sunday, March 26, 2017 9:28 PM  
**To:** Bell, Danielle; Cameron, John IV; Conrad Juergensmeyer; Hutchinson, Harold; Ivette Miranda; Bell, Danielle; Page, Mike; Paula Tyrell; Sonya Wolthuis; Stevens, Teresa  
**Subject:** March 2017 Agency Inspections  
**Attachments:** March 2017 Agency Inspections 03-15-2017.pdf; March 2107 Agency Inspections 03-26-2017.pdf

**Follow Up Flag:** Follow up  
**Flag Status:** Flagged

Good Evening,

Attached are the March 2017 Agency Inspections for the Hernando County Sheriff's Office.

Instrument 814 and 5249 were not inspected due to both of them being at CMI for servicing.

Instrument 228 is an Evidential Loaner instrument that we have received from FDLE while 814 and 5249 are being serviced.

Instrument 812 and 228 were initially inspected on 03/15/2017. Both instruments passed inspections but were subject to RFI detects. During the month of March, we were having a large sliding door replaced in our Booking department. There were some malfunctions during the installation process that delayed the installation. Once the installation was completed, I ran 20 stability checks through each instruments with a .20% Alcohol Reference Solution. There were no RFI detect errors during the stability checks. Additionally, there have not been any RFI detects since the installation of the sliding door was completed.

Sgt. Stevens has received formal quotes for the installation of panels that will assist in blocking any radio frequency interference from the Breath Test Unit.

I conducted an additional agency inspection this evening and both instruments passed without any discrepancies. I have attached the initial agency inspection from 03/15/2017, the stability checks for both instruments, and the agency inspection from 03/26/2017.

If I can be of additional assistance, please do not hesitate to contact me.

Respectfully,  
Sgt. B. Collito

Sergeant Brad Collito  
 Intake/ Release Unit  
 Hernando County Detention Center  
 16425 Spring Hill Dr  
 Brooksville, FL 34604  
 T(352)544-2334 F(352)544-2349  
[bcollito@hernandosheriff.org](mailto:bcollito@hernandosheriff.org)

This procedure was discussed with me & determined that the repeated RFI detects were a result of external factors & not the instrument & those have been resolved.

AB

80-000812  
80-000814  
80-005249

ND

80-000228

**Bell, Danielle**

**From:** Collito, Brad <bcollito@hernandosheriff.org>  
**Sent:** Wednesday, February 15, 2017 12:52 PM  
**To:** Cameron, John IV; Conrad Juergensmeyer; Hutchinson, Harold; Ivette Miranda; Bell, Danielle; Page, Mike; Paula Tyrell; Sonya Wolthuis; Stevens, Teresa  
**Subject:** Hernando County Sheriff's Office February 2017 Agency Inspections  
**Attachments:** February 2017 Agency Inspections.pdf

Good Afternoon,

Enclosed are the February 2017 Agency Inspections for the Hernando County Sheriff's Office.

Instrument 812 passed the inspection with no discrepancies.

Instrument 814 passed the inspection with no discrepancies and is being sent to FDLE for its annual inspection and to have the EEPROM checked due to an initial failure during the diagnostic when coming on from being completely powered off.

Instrument 5249 is currently at FDLE and was not inspected.

We have received instrument 228 from FDLE as an evidential loaner instrument due to two of our instruments being at FDLE. I have signed the Equipment Loan Agreement and attached it with this month's documentation. An agency inspection was completed on 228 with no discrepancies.

If you have any questions, please do not hesitate to contact me,

Thank you,

Sergeant Brad Collito  
Intake/ Release Unit  
Hernando County Detention Center  
16425 Spring Hill Dr  
Brooksville, FL 34604  
T(352)544-2334 F(352)544-2349

[bcollito@hernandosheriff.org](mailto:bcollito@hernandosheriff.org)<<mailto:bcollito@hernandosheriff.org>>

Please note: Florida has a very broad Public Records Law. Most written communications to or from the Hernando County Sheriff's Office are public records available to anyone upon request. Your e-mail, including your e-mail address, may therefore, be subject to public disclosure.