

80-000775
②

Shanahan, Jake

From: Brett Sullivan <Brett.Sullivan@columbiasheriff.org>
Sent: Monday, January 16, 2017 5:48 AM
To: Shanahan, Jake
Subject: CCSO Agency Inspection

Jake,

I conducted an agency inspection on Intoxilyzer 80-000775 this morning, several actually. The control test for the .20 was reading low, a few times under .190 but mostly between .190-.193. I conducted three inspections marking the instrument as not complying, due to a mental error of forgetting the acceptable range for the .20 test. I forgot the acceptable range and thought it was .195-.205, causing me to do several unnecessary inspections. I conducted a fourth, correctly remembered the acceptable range and marked the instrument as in compliance. I wanted to let you know what happened and make you aware of the incident. I sincerely apologize for this mistake.

Thank you,
D/S Brett Sullivan
Columbia County Sheriff's Office
4917 US Highway 90 E
Lake City, FL 32055



THIS E-MAIL, ITS CONTENT AND ANY FILES TRANSMITTED WITH IT ARE CONFIDENTIAL AND ARE INTENDED SOLELY FOR THE USE OF THE INDIVIDUAL OR ENTITY TO WHOM THEY ARE ADDRESSED. IF YOU ARE NOT THE INTENDED RECIPIENT OR THE PERSON RESPONSIBLE FOR DELIVERING THE E-MAIL TO THE INTENDED RECIPIENT. BE ADVISED THAT YOU HAVE RECEIVED THIS E-MAIL IN ERROR AND THAT ANY USE, COPYING, PRINTING, FORWARDING OR DISSEMINATION OF THIS E-MAIL IS STRICTLY PROHIBITED. IF YOU HAVE RECEIVED THIS E-MAIL IN ERROR, PLEASE IMMEDIATELY NOTIFY THE SENDER OF THIS E-MAIL BY REPLY OR CALL (386) 758-1378 AND DELETE IT FROM YOUR SYSTEM IMMEDIATELY.