

80-000900

Murphy, Patrick

From: Johnston, Robert <Robert.Johnston@MyFWC.com>
Sent: Thursday, February 02, 2017 1:54 PM
To: Murphy, Patrick
Subject: Intoxilyzer 80-000900

Patrick,

As per our previous conversation, Intoxilyzer ser# 80-000900 failed it's inspection on November 29, 2016. The cause for the failure was mouth alcohol in the tube causing Ambient Fail. The instrument has been out of service since November 29, 2016. I request permission to inspect the instrument again and put the instrument back in service upon passing inspection.

Thank you,

Officer Robert Johnston M809 FWC