

863-644-~~3325~~ 3325

**Return Material Authorization form**

**SHIP TO:**

CMI, Inc.

Enforcement Electronics

Shipment to repair facility authorized by: Michael Weiskopf on 9/6/17

**Items Returned:**

Instrument  Supplies  other  Please Describe: \_\_\_\_\_

Instrument Model: Intoxilyzer 8000 Serial Number 80-001051

Bill To Address:  
St. Petersburg Police Dept.  
1300 1st Ave N.  
St. Petersburg, FL  
33705

Ship To Address:  
FDLE-Danielle Bell  
2331 Phillips Rd.  
Tallahassee, FL  
32308

**Reason for Return:**

D/C fuse on back of instrument spun instead of the fuse  
box staying in place => the fuse itself unseating. D/C fuse box  
=> fuse combo spun => connector broke. Instrument working on  
A/C power just fine.

**Please choose one of the following options:**

- 1. I \_\_\_\_\_, authorize all repairs.
- 2. I \_\_\_\_\_, authorize repairs up to \$ \_\_\_\_\_.
- 3. I require an estimate **BEFORE** any repairs will be authorized and/or conducted,  
Please contact, Name: Michael Weiskopf Phone Number: 727 893 7099

**Notes:**