

Bell, Danielle

From: Wiegert, Kimbra <KWiegert@polksheriff.org>
Sent: Wednesday, December 27, 2017 7:51 PM
To: Bell, Danielle
Cc: Pickles, Tamara
Subject: 1127

Follow Up Flag: Follow up
Flag Status: Flagged

During the agency inspection, 1127 would not come online, the display showed just the intoxilyzer 8000, cmi information. I restarted it and tried another cord with no results. Per our discussion I am sending it to you to check the instrument out. I do not believe that the instrument was used to perform any tests since the last agency inspection in November.

Kimbra L. Wiegert #4591 BA/CJS, CPM

Sergeant.

Department of Detention

Inmate Intake Section

Polk County Sheriff's Office

863-534-6300-(Office)

863-534-6343(Fax)

Bell, Danielle

From: Wilson, Barbie <BWilson@polksheriff.org>
Sent: Thursday, September 07, 2017 8:51 AM
To: Bell, Danielle
Subject: RE: Afternoon

ok, thank you and have a good day

Detention Deputy Barbie Wilson, 6493
Inmate Intake Trusty Supervisor
Agency Inspector

Office: 863-534-6248 (ext. 7919)
Cell: (863)287-9128

"One measure of your success will be the degree to which you build up others who work with you. While building up others, you will build up yourself."

— James E. Casey

From: Bell, Danielle [DanielleBell@fdle.state.fl.us]
Sent: Thursday, September 7, 2017 8:36 AM
To: Wilson, Barbie
Cc: Public Records; Lawson-Colston, Barbara
Subject: RE: Afternoon

Here is the Form 40 for 1127 as requested. Also, I'll put Mouth Alcohol in the mail today, but no guarantees on when it makes it down to you (considering the storm.) I'll check out 1237 when I get it and let you know what I find.

Thanks,

Danielle

Danielle M. Bell
Department Inspector
Alcohol Testing Program
Florida Department of Law Enforcement
(850) 617-1286 (office)
(850) 274-2757 (cell)
(850) 921-3787 (fax)

-----Original Message-----

From: Wilson, Barbie [mailto:BWilson@polksheriff.org]
Sent: Wednesday, September 06, 2017 2:46 PM
To: Bell, Danielle
Subject: Afternoon

Afternoon,

I was hoping for a good day today but unfortunately Instrument 80-001237 did not cooperate. I conducted 3 different test and it did not pass the Interferent Detect test on either inspection. I checked the to ensure that it was at the correct temperature, it appeared to be at 33 degrees. I waited another 10-15 minutes for it to heat up even more and this did not solve the problem. On the third attempt, I switched simulators and again it did not pass. I have printed and uploaded the test results and have removed it from use.

Do you want me to send it to you or to Enforcement Electronics? Also, could you please send me a copy of FDLE Form 40 for Instrument 80-001127?

Thank you for your assistance and stay safe during the storm.

Detention Deputy Barbie Wilson, 6493
Inmate Intake Trusty Supervisor
Agency Inspector

Office: 863-534-6248 (ext. 7919)
Cell: (863)287-9128

"One measure of your success will be the degree to which you build up others who work with you. While building up others, you will build up yourself."

— James E. Casey

From: Bell, Danielle [DanielleBell@fdle.state.fl.us]
Sent: Friday, September 1, 2017 12:34 PM
To: Wilson, Barbie
Subject: Re: Form-40-Agency-Insp-Rpt-Intox-80001

Perfect. If you'll just sign it. That's great.

Thanks,
Danielle

Sent from my Verizon, Samsung Galaxy smartphone

----- Original message -----

From: "Wilson, Barbie" <BWilson@polksheriff.org>
Date: 9/1/17 12:31 PM (GMT-05:00)
To: "Bell, Danielle" <DanielleBell@fdle.state.fl.us>, "Wilson, Barbie" <BWilson@polksheriff.org>
Subject: Form-40-Agency-Insp-Rpt-Intox-80001

Bell, Danielle

From: Wilson, Barbie <BWilson@polksheriff.org>
Sent: Tuesday, May 09, 2017 4:07 PM
To: Bell, Danielle
Subject: FW: Scan Job from Copier
Attachments: scan.pdf

80-001127
80-001237
80-001080

I truly had an off day today conducting my inspections. I have indicated why I received the errors on the test, here is a copy of all three test. Have you received BA instrument from Enforcement Electronics belonging to the Polk County Sheriff's Office? I will be out of the office until tomorrow morning at 7:30 if I need to explain further and have court tomorrow afternoon.

Wilson

From: copier@polksheriff.org [copier@polksheriff.org]
Sent: Tuesday, May 9, 2017 3:17 PM
To: Wilson, Barbie
Subject: Scan Job from Copier

Bell, Danielle

From: Bell, Danielle
Sent: Thursday, February 09, 2017 1:19 PM
To: 'Wilson, Barbie'
Cc: Kirkland, Brett
Subject: RE: Scan Job from Copier

Importance: High

Follow Up Flag: Follow up
Due By: Tuesday, February 14, 2017 10:30 AM
Flag Status: Completed

Tracking:	Recipient	Read
	'Wilson, Barbie'	
	Kirkland, Brett	Read: 2/9/2017 1:21 PM

After some discussion, in an overabundance of caution, please send Instrument #80-001127 to me before any additional breath tests are conducted on it. If any breath tests have been conducted on the instrument since the inspection today at 10:28, please do another inspection prior to sending. Otherwise, please simply box the instrument appropriately and send it to me for further testing as well as its annual Department Inspection.

Thanks,

Danielle

Danielle M. Bell
 Department Inspector
 Alcohol Testing Program
 Florida Department of Law Enforcement
 (850) 617-1286 (office)
 (850) 274-2757 (cell)
 (850) 921-3787 (fax)

-----Original Message-----

From: Wilson, Barbie [<mailto:BWilson@polksheriff.org>]
Sent: Thursday, February 09, 2017 12:20 PM
To: Bell, Danielle
Subject: RE: Scan Job from Copier

I didn't do anything to it. The conditions were the absolute same, I just conducted another test and it passed. This instrument has been being used alot lately so I just switched it with another instrument and it is now being used for back up. It's not actually being removed just not being used as the first instrument at this time.

From: Bell, Danielle [DanielleBell@fdle.state.fl.us]
Sent: Thursday, February 9, 2017 11:37 AM
To: Wilson, Barbie

Subject: RE: Scan Job from Copier

Good morning,

As we have discussed in the past, please indicate on the Form 40 for instrument 80-001127 what was done prior to repeating the test that brought the instrument into compliance. Was the change external to the instrument (ie: the simulator required more time to heat?) Once the notation has been made, please initial and date the change and send me a copy. Also, you indicated that you took it out of service; is your plan to send it to me for its annual inspection or do you plan to do another Agency Inspection and place it back in service at a later date? With regard to 80-001055, I did not know it was at Enforcement Electronics. The last information I received regarding 80-001055 was on August 5, 2016 indicating you preferred that I send it back to your agency. I have not seen it since that time.

Thank you for the communication,

Danielle

Danielle M. Bell
Department Inspector
Alcohol Testing Program
Florida Department of Law Enforcement
(850) 617-1286 (office)
(850) 274-2757 (cell)
(850) 921-3787 (fax)

-----Original Message-----

From: Wilson, Barbie [<mailto:BWilson@polksheriff.org>]
Sent: Thursday, February 09, 2017 11:26 AM
To: Bell, Danielle
Subject: FW: Scan Job from Copier

Good Morning,

I just wanted to advise you that Instrument 80-001127 failed the .20 inspection the first time and the error message was Control Outside Tolerance; however, after performing the test the second time it passed. I took it out of service it any way and replaced it with another instrument for use. Also, did you receive Instrument 80-001055 from Enforcement Electronics because I never received it back from having it serviced. I will give them a call today.

Thank you,

Wilson