

Bell, Danielle

From: Michael D. Weiskopf <Michael.Weiskopf@stpete.org>
Sent: Sunday, January 15, 2017 12:54 AM
To: Bell, Danielle
Subject: Intoxilyzer unit 80-001653

I'm writing this as a follow-up to our telephone conversation. While conducting the agency inspection on instrument 80-001653, it was found to be not in compliance while testing the .20 solution. I instrument read .190, .189, .189. I took the instrument out of service and contacted you, leaving a voice mail message. Upon your return call and troubleshooting the problem, it was determined that the issue was with the simulator not heating properly and not with the Intoxilyzer. I re-inspected the instrument and it passed without issue.

Thank you for your help in this matter.

Officer Michael Weiskopf
DUI Unit/Agency Inspector
St Petersburg Police Department
1300 1st Avenue North
St Petersburg, FL 33705
Office: 727-893-7099
Audix: 727-551-3108

Your Sunshine City