

**Bell, Danielle**

---

**From:** Art Rowand <ARowand@pascosheriff.org>  
**Sent:** Wednesday, July 26, 2017 2:44 PM  
**To:** Bell, Danielle  
**Subject:** Monthly Inspections

Today, during the July monthly inspection the first bottle of .05 solution was bad. I ran the test and it showed .000. I emptied that bottle and used another bottle and retested and it passed.

The instruments affected were 80-001074 & 80-003347.

The first .05 bottle (Bad One) was:

Lot: 201603D  
Exp: 03/08/2018  
Bottle #: 0353

Second .05 Bottle

Lot: 201603D  
Exp: 03/08/2018  
Bottle #: 0325

With doing mouth alcohol on Instrument 80-001081, the first attempt did not show slope Not Met instead read the mouth alcohol. The second attempt read Slope Not Met and passed.

If you need any additional please advise.

**Art Rowand, Sergeant**  
**Pasco Sheriff's Office**  
**Special Operations Division**  
**Traffic Enforcement Section**

**727-844-7798**

[arowand@pascosheriff.org](mailto:arowand@pascosheriff.org)

This correspondence is from a law enforcement agency and may contain confidential, classified, and/or privileged communications. Unauthorized use of this information may subject you to criminal and civil penalties. This document is the property of the Pasco Sheriff's Office and may not be disseminated without authorization. As such this document may be exempt from public record as defined in Florida State Statute Chapter 119