

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

SEP 19 2017

FDLE

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: COLUMBIA COUNTY SO
Time of Inspection: 12:04

Date of Inspection: 09/19/2017

Serial Number: 80-000775
Software: 8100.27

VAC

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG619605 Exp: 07/14/2018
0.000	0.049	0.080	0.198	0.081
0.000	0.049	0.081	0.200	0.080
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.080	0.199	0.080
0.000	0.049	0.080	0.199	0.080
0.000	0.049	0.081	0.199	0.081
0.000	0.049	0.080	0.199	0.080
0.000	0.049	0.081	0.199	0.080
0.000	0.049	0.081	0.199	0.080
0.000	0.050	0.081	0.198	0.081

Standard Deviations	0.0003	0.0005	0.0006	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

Jam

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAKE L SHANAHAN

Signature and Printed Name

09/19/2017
Date

*9/19/17
JD*