

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

OCT 11 2017

FDLE

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: NAS JACKSONVILLE PD
Time of Inspection: 09:34

Date of Inspection: 10/11/2017

Serial Number: 80-000778
Software: 8100.27

low

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201707D Exp: 07/25/2019	0.08g/210L Test (g/210L) Lot#:201707E Exp: 07/25/2019	0.20g/210L Test (g/210L) Lot#:201707C Exp: 07/24/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG708807 Exp: 03/29/2019
/ 0.000	0.050	0.080	0.202	0.078
/ 0.000	0.049	0.082	0.206	0.078
/ 0.000	0.049	0.080	0.206	0.079
/ 0.000	0.050	0.080	0.206	0.078
/ 0.000	0.050	0.081	0.204	0.078
/ 0.000	0.049	0.082	0.205	0.079
/ 0.000	0.050	0.081	0.203	0.078
/ 0.000	0.050	0.081	0.204	0.078
/ 0.000	0.051	0.081	0.205	0.079
/ 0.000	0.050	0.081	0.204	0.078

Standard Deviations	0.0006	0.0007	0.0013	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0007 Number of Simulators Used: 5

Remarks:

00: Ambient Fail-MA VAPORS PRESENT.

SP

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAKE L SHANAHAN

Signature and Printed Name

10/11/2017
Date

*10/11/17
JL*